



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111289

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/02/2020	13:35
2	L4200-02		1	8	100.00	>212.0	>212.0	10/02/2020	14:10
3	L4200-02DUP		1	8	100.00	>212.0	>212.0	10/02/2020	14:40
4	L4232-02		1	9	100.00	>212.0	>212.0	10/02/2020	15:15
5	L4233-01		1	8	100.00	>212.0	>212.0	10/02/2020	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111289

WorkList Name : fp10120 WorkList ID : 142844 Department : Wet-Chemistry Date : 10-01-2020 09:17:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/13/2020	Water	L4200-02 J	Flash Point	Cool 4 deg C	ARCA05	O12	PB-2020-WW-000-01	09/29/2020	1010A
10/08/2020	Water	L4232-02 B	Flash Point	Cool 4 deg C	PSEG01	O21	1613	10/01/2020	1010A
10/15/2020	Water	L4233-01 H	Flash Point	Cool 4 deg C	DIST01		55-ST-29	10/01/2020	1010A

Date/Time 10/2/2020 13:05
 Received by: KS (wehch)
 Relinquished by: KS (wehch)

Date/Time 10/2/2020 15:50
 Received by: KS (wehch)
 Relinquished by: KS (wehch)