



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111300

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/03/2020	11:30
2	CAL2	1	Water	NA	NA	20.1	7.00	10/03/2020	11:31
3	CAL3	1	Water	NA	NA	20.2	10.03	10/03/2020	11:32
4	ICV	1	Water	NA	NA	20.1	6.98	10/03/2020	11:35
5	CCV1	1	Water	NA	NA	20.2	2.02	10/03/2020	11:36
6	L4260-04	1	Solid	20.03	20	20.3	8.61	10/03/2020	11:38
7	L4260-06	1	Solid	20.04	20	20.4	8.52	10/03/2020	11:39
8	L4260-12	1	Solid	20.01	20	20.2	8.60	10/03/2020	11:40
9	L4260-14	1	Solid	20.03	20	20.3	8.19	10/03/2020	11:42
10	L4264-01	1	Solid	20.04	20	20.2	8.42	10/03/2020	11:45
11	L4264-01DUP	1	Solid	20.05	20	20.3	8.43	10/03/2020	11:46
12	CCV2	1	Water	NA	NA	20.1	12.03	10/03/2020	11:48

2B11300

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L4264 WorkList ID : 142919 Department : Wet-Chemistry Date : 10-03-2020 10:52:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2020	Solid	L4260-04	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B21-100220-0-10	10/02/2020	9045D
10/02/2020	Solid	L4260-06	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B21-100220-10-18.5	10/02/2020	9045D
10/02/2020	Solid	L4260-12	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B23-100220-0-10	10/02/2020	9045D
10/02/2020	Solid	L4260-14	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B23-100220-10-19	10/02/2020	9045D
10/02/2020	Solid	L4264-01	Corrosivity	Cool 4 deg C	URSC01	O11	BPC-WASTE-DRUM	10/02/2020	9045D

Date/Time 10-03-2020 11:00
 Received by: JP (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 10-03-2020 15:00
 Received by: KS (wet-chem)
 Relinquished by: JP (wet-chem)