


LB111301

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.065	-0.065	555	0.000	555				
0.2	0.200	0.203	0.003	1461	0.0 %	1461				
2.5	2.500	2.522	0.022	9312	0.0 %	9312				
5.0	5.000	5.192	0.192	18351	0.0 %	18351				
7.5	7.500	7.327	-0.173	25578	0.0 %	25578				
10.0	10.000	10.021	0.021	34695	0.0 %	34695				

LB111301 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Oct 2020 08:57:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	555	-	0.00007470A	PPB		05 Oct 2020 09:07:04	S	Std
0.2 - 1	50.2	1461	-	0.20007470A	PPB		05 Oct 2020 09:09:21	S	Std
2.5 - 1	52.5	9312	-	2.50007470A	PPB		05 Oct 2020 09:11:39	S	Std
5.0 - 1	55	18351	-	5.00007470A	PPB		05 Oct 2020 09:13:55	S	Std
7.5 - 1	57.5	25578	-	7.50007470A	PPB		05 Oct 2020 09:16:12	S	Std
10.0 - 1	50	34695	-	10.00007470A	PPB		05 Oct 2020 09:18:28	S	Std
ICV41 - 1	ICV41	14176	3.9589	-7470A	PPB		05 Oct 2020 09:21:18	U	SMPL
ICB41 - 1	ICB41	837	0.0184	-7470A	PPB		05 Oct 2020 09:33:36	U	SMPL
CCV49 - 1	CCV49	17137	4.8337	-7470A	PPB		05 Oct 2020 09:35:51	U	SMPL
CCB49 - 1	CCB49	731	-0.0130	-7470A	PPB		05 Oct 2020 09:38:06	U	SMPL
CRA - 1	CRA	1357	0.1720	-7470A	PPB		05 Oct 2020 09:40:21	U	SMPL
HighStd - 1	HighStd	32505	9.3737	-7470A	PPB		05 Oct 2020 09:42:36	U	SMPL
ChkStd - 1	ChkStd	22177	6.3226	-7470A	PPB		05 Oct 2020 09:44:52	U	SMPL
PB132071BL - 1	PBW	692	-0.0245	-7470A	PPB		05 Oct 2020 09:49:43	U	SMPL
PB132071BS - 1	LCSW	12193	3.3731	-7470A	PPB		05 Oct 2020 09:51:58	U	SMPL
L3865-04 - 1	NB-08-093020	515	-0.0768	-7470A	PPB		05 Oct 2020 09:54:15	U	SMPL
L3865-04DUP - 1	NB-08-093020DUP	559	-0.0638	-7470A	PPB		05 Oct 2020 09:56:31	U	SMPL
L3865-04MS - 1	NB-08-093020MS	12781	3.5468	-7470A	PPB		05 Oct 2020 09:58:48	U	SMPL
L3865-04MSD - 1	NB-08-093020MSD	13362	3.7185	-7470A	PPB		05 Oct 2020 10:01:05	U	SMPL
L3865-08 - 1	NB-07-093020	502	-0.0806	-7470A	PPB		05 Oct 2020 10:03:20	U	SMPL
CCV50 - 1	CCV50	16958	4.7808	-7470A	PPB		05 Oct 2020 10:05:36	U	SMPL
CCB50 - 1	CCB50	693	-0.0242	-7470A	PPB		05 Oct 2020 10:07:51	U	SMPL
L3872-16 - 1	BU-03-093020	597	-0.0525	-7470A	PPB		05 Oct 2020 10:10:07	U	SMPL
L4198-02 - 1	RO-15	660	-0.0339	-7470A	PPB		05 Oct 2020 10:12:23	U	SMPL
L4198-05 - 1	RO-21	674	-0.0298	-7470A	PPB		05 Oct 2020 10:14:38	U	SMPL
L4198-08 - 1	RO-25	678	-0.0286	-7470A	PPB		05 Oct 2020 10:16:53	U	SMPL
L4198-11 - 1	RO-24	652	-0.0363	-7470A	PPB		05 Oct 2020 10:19:09	U	SMPL
L4198-14 - 1	RO-10	627	-0.0437	-7470A	PPB		05 Oct 2020 10:21:26	U	SMPL
L4198-17 - 1	RO-5	700	-0.0221	-7470A	PPB		05 Oct 2020 10:23:42	U	SMPL
L4198-20 - 1	RO-20	668	-0.0316	-7470A	PPB		05 Oct 2020 10:26:00	U	SMPL
L4198-23 - 1	RO-17	681	-0.0277	-7470A	PPB		05 Oct 2020 10:28:16	U	SMPL
L4198-29 - 1	RO-19	686	-0.0262	-7470A	PPB		05 Oct 2020 10:30:32	U	SMPL
CCV51 - 1	CCV51	17709	5.0027	-7470A	PPB		05 Oct 2020 10:32:48	U	SMPL
CCB51 - 1	CCB51	750	-0.0073	-7470A	PPB		05 Oct 2020 10:45:27	U	SMPL
L4198-32 - 1	RO-4	682	-0.0274	-7470A	PPB		05 Oct 2020 10:47:42	U	SMPL
L4199-02 - 1	11940	676	-0.0292	-7470A	PPB		05 Oct 2020 10:49:58	U	SMPL
PB132072BL - 1	PBW	745	-0.0088	-7470A	PPB		05 Oct 2020 10:52:13	U	SMPL
PB132072BS - 1	LCSW	12610	3.4963	-7470A	PPB		05 Oct 2020 10:54:29	U	SMPL
L4200-02 - 1	PB-2020-WW-000-01	879	0.0308	-7470A	PPB		05 Oct 2020 10:56:45	U	SMPL
L4200-02DUP - 1	PB-2020-WW-000-01DUP	916	0.0417	-7470A	PPB		05 Oct 2020 10:59:01	U	SMPL
L4200-02MS - 1	PB-2020-WW-000-01MS	16678	4.6981	-7470A	PPB		05 Oct 2020 11:01:17	U	SMPL
L4200-02MSD - 1	PB-2020-WW-000-01MSD	15554	4.3660	-7470A	PPB		05 Oct 2020 11:03:34	U	SMPL
L4232-01 - 1	1613	1489	0.2110	-7470A	PPB		05 Oct 2020 11:05:51	U	SMPL
L4242-21 - 1	FIELD-BLANK	606	-0.0499	-7470A	PPB		05 Oct 2020 11:08:06	U	SMPL
CCV52 - 1	CCV52	18046	5.1022	-7470A	PPB		05 Oct 2020 11:10:22	U	SMPL
CCB52 - 1	CCB52	721	-0.0159	-7470A	PPB		05 Oct 2020 11:15:22	U	SMPL
PB132071TB - 1	PB132029TB	653	-0.0360	-7470A	PPB		05 Oct 2020 11:17:37	U	SMPL
L3865-04LX5 - 1		712	-0.0186	-7470A	PPB		05 Oct 2020 11:19:53	U	SMPL
L3865-04A - 1		13634	3.7988	-7470A	PPB		05 Oct 2020 11:22:08	U	SMPL
L4200-02LX5 - 1		653	-0.0360	-7470A	PPB		05 Oct 2020 11:24:24	U	SMPL
L4200-02A - 1		15573	4.3716	-7470A	PPB		05 Oct 2020 11:26:40	U	SMPL
CCV53 - 1	CCV53	17585	4.9660	-7470A	PPB		05 Oct 2020 11:28:56	U	SMPL
CCB53 - 1	CCB53	845	0.0207	-7470A	PPB		05 Oct 2020 11:36:07	U	SMPL