



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/6/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:05
In Date: 10/05/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 10/06/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111318

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L4269-01	362-CONCRETE	22	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4269-02	362-CONCRETE	23	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4271-01	GV-CONCRETE	24	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4273-01	SB-1	1	1.12	9.63	8.38	85.3	
L4273-02	SB-1A	2	1.14	9.9	9.09	90.8	
L4273-03	SB-2	3	1.16	9.65	8.48	86.2	
L4273-04	SB-2A	4	1.13	9.6	8.5	87.0	
L4273-05	SB-3	5	1.16	9.9	7.88	76.9	
L4273-06	SB-3A	6	1.15	9.95	9.12	90.6	
L4273-07	SB-4	7	1.13	9.81	8.53	85.3	
L4273-08	SB-4A	8	1.15	9.98	9.06	89.6	
L4273-09	SB-5	9	1.11	9.55	7.55	76.3	
L4273-10	SB-5A	10	1.14	9.55	8.51	87.6	
L4273-11	SB-6	11	1.18	9.75	8.63	86.9	
L4273-12	SB-6A	12	1.14	9.96	9.1	90.2	
L4273-13	SB-7	13	1.11	9.97	8.51	83.5	
L4273-14	SB-7A	14	1.1	9.81	8.65	86.7	
L4273-15	SB-8	15	1.14	9.96	8.69	85.6	
L4273-16	SB-8A	16	1.13	9.55	8.5	87.5	
L4273-17	SB-9	17	1.13	9.51	7.86	80.3	
L4273-18	SB-9A	18	1.13	9.68	8.37	84.7	
L4273-19	SB-10	19	1.19	9.61	8.35	85.0	
L4273-20	SB-10A	20	1.18	9.93	8.93	88.6	
L4274-01	WC-1	21	1.13	9.77	8.66	87.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$