



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111322

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/05/2020	16:40
2	CAL2	1	Water	NA	NA	20.2	7.00	10/05/2020	16:41
3	CAL3	1	Water	NA	NA	20.2	10.03	10/05/2020	16:42
4	ICV	1	Water	NA	NA	20.2	7.01	10/05/2020	16:44
5	CCV1	1	Water	NA	NA	20.3	2.02	10/05/2020	16:45
6	L4270-01	1	Water	NA	NA	20.1	7.31	10/05/2020	16:46
7	L4270-01DUP	1	Water	NA	NA	20.2	7.32	10/05/2020	16:47
8	L4270-03	1	Water	NA	NA	19.8	7.26	10/05/2020	16:49
9	L4270-05	1	Water	NA	NA	19.9	7.40	10/05/2020	16:51
10	L4270-07	1	Water	NA	NA	22.0	7.15	10/05/2020	17:00
11	CCV2	1	Water	NA	NA	20.2	12.02	10/05/2020	17:02

2B111322

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4270 WorkList ID : 142965 Department : Wet-Chemistry Date : 10-05-2020 16:32:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2020	Water	L4270-01	-A pH	Cool 4 deg C	PSEG01	O23	TOTE-COMP-1	10/05/2020	9040C
10/05/2020	Water	L4270-03	-A pH	Cool 4 deg C	PSEG01	O23	TOTE-COMP-2	10/05/2020	9040C
10/05/2020	Water	L4270-05	-A pH	Cool 4 deg C	PSEG01	O23	TOTE-COMP-3	10/05/2020	9040C
10/05/2020	Water	L4270-07	-A pH	Cool 4 deg C	PSEG01	O23	TOTE-COMP-4	10/05/2020	9040C

Date/Time 10.05.2020 16:35
Received by: Jp/boet et al
Relinquished by: Jp/boet et al

Date/Time 10.05.2020 17:35
Received by: Jp/boet
Relinquished by: Jp/boet