



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111344

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/07/2020	09:20
2	CAL2	1	Water	NA	NA	20.1	7.00	10/07/2020	09:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/07/2020	09:22
4	ICV	1	Water	NA	NA	20.2	7.00	10/07/2020	09:23
5	CCV1	1	Water	NA	NA	20.1	2.01	10/07/2020	09:25
6	L4280-02	1	Water	NA	NA	20.1	6.88	10/07/2020	09:26
7	L4280-02DUP	1	Water	NA	NA	20.2	6.89	10/07/2020	09:27
8	CCV2	1	Water	NA	NA	20.1	12.02	10/07/2020	09:30

RB 1134h

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COROSIVITY L4280 WorkList ID : 143032 Department : Wet-Chemistry Date : 10-07-2020 09:12:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/06/2020	Water	L4280-02	-B Corrosivity	Cool 4 deg C	PSEG01	O31	10-6-HSFRAC	10/06/2020	9040C

Date/Time 10-07-2020 9:15
 Received by: JA / [Signature]
 Relinquished by: JA / [Signature]

Date/Time 10-07-2020 10:15
 Received by: JA / [Signature]
 Relinquished by: JA / [Signature]