



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111348

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/07/2020	09:05
2	L4270-01		1	8	100.00	>212.0	>212.0	10/07/2020	09:35
3	L4270-01DUP		1	8	100.00	>212.0	>212.0	10/07/2020	09:58
4	L4270-03		1	9	100.00	>212.0	>212.0	10/07/2020	10:20
5	L4270-05		1	9	100.00	>212.0	>212.0	10/07/2020	10:50
6	L4270-07		1	8	100.00	>212.0	>212.0	10/07/2020	11:10
7	L4280-02		1	9	100.00	>212.0	>212.0	10/07/2020	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP10620 WorkList ID : 143011 Department : Wet-Chemistry Date : 10-06-2020 17:17:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/07/2020	Water	L4270-01 B	Flash Point	Cool 4 deg C	PSEG01	O23	TOTE-COMP-1	10/05/2020	1010A
10/07/2020	Water	L4270-03 B	Flash Point	Cool 4 deg C	PSEG01	O23	TOTE-COMP-2	10/05/2020	1010A
10/07/2020	Water	L4270-05 B	Flash Point	Cool 4 deg C	PSEG01	O23	TOTE-COMP-3	10/05/2020	1010A
10/07/2020	Water	L4270-07 B	Flash Point	Cool 4 deg C	PSEG01	O23	TOTE-COMP-4	10/05/2020	1010A
10/09/2020	Water	L4280-02 C	Flash Point	Cool 4 deg C	PSEG01	O31	10-6-HSFRAC	10/06/2020	1010A

Date/Time 10/7/2020 08:30
 Received by: KS (webhu)
 Relinquished by: Fe (ext)

Date/Time 10/7/2020 11:45
 Received by: Fe (ext lab)
 Relinquished by: KS (webhu)