



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111352

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/07/2020	13:10
2	CAL2	1	Water	NA	NA	20.1	6.98	10/07/2020	13:11
3	CAL3	1	Water	NA	NA	20.1	10.03	10/07/2020	13:13
4	ICV	1	Water	NA	NA	20.2	7.00	10/07/2020	13:15
5	CCV1	1	Water	NA	NA	20.1	2.02	10/07/2020	13:16
6	L4301-01	1	Water	NA	NA	16.6	6.45	10/07/2020	13:19
7	L4301-01DUP	1	Water	NA	NA	16.7	6.46	10/07/2020	13:20
8	L4301-03	1	Water	NA	NA	17.6	6.54	10/07/2020	13:22
9	CCV2	1	Water	NA	NA	20.1	12.02	10/07/2020	13:24

LB 111352

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14301

WorkList ID : 143047

Department : Wet-Chemistry

Date : 10-07-2020 12:55:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/06/2020	Water	L4301-01	pH	Cool 4 deg C	APT101	O23	WGGM-20-WW-TSF-033	10/06/2020	SM 4500-H B
10/06/2020	Water	L4301-03	pH	Cool 4 deg C	APT101	O23	WGGM-20-WW-TSF-034	10/06/2020	SM 4500-H B

Date/Time 10-07-2020 13:00
Received by: Jd / wet
Relinquished by: Jd / wet

Date/Time 10-07-2020 16:00
Received by: Jd / wet
Relinquished by: Jd / wet