



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111368

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/07/2020	16:35
2	CAL2	1	Water	NA	NA	20.2	7.00	10/07/2020	16:36
3	CAL3	1	Water	NA	NA	20.2	10.03	10/07/2020	16:37
4	ICV	1	Water	NA	NA	20.2	6.98	10/07/2020	16:39
5	CCV1	1	Water	NA	NA	20.2	2.01	10/07/2020	16:40
6	L4291-04	1	Solid	20.03	20	22.6	9.60	10/07/2020	17:15
7	L4291-06	1	Solid	20.04	20	22.4	9.37	10/07/2020	17:16
8	L4291-12	1	Solid	20.02	20	22.4	9.26	10/07/2020	17:18
9	L4292-04	1	Solid	20.03	20	22.6	8.45	10/07/2020	17:20
10	L4292-10	1	Solid	20.04	20	22.5	8.23	10/07/2020	17:22
11	L4311-04	1	Solid	20.01	20	22.5	7.72	10/07/2020	17:25
12	L4311-06	1	Solid	20.03	20	22.7	7.42	10/07/2020	17:26
13	L4311-06DUP	1	Solid	20.04	20	22.8	7.43	10/07/2020	17:27
14	CCV2	1	Water	NA	NA	20.2	12.02	10/07/2020	17:28

WORKLIST(Hardcopy Internal Chain)

111368

WorkList Name : CORROSIVITY L4291 WorkList ID : 143055 Department : Wet-Chemistry Date : 10-07-2020 15:39:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/06/2020	Solid	L4291-04	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B4-100620-0-10	10/06/2020	9045D
10/06/2020	Solid	L4291-06	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B4-100620-10-20	10/06/2020	9045D
10/06/2020	Solid	L4291-12	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B4-100620-DP	10/06/2020	9045D
10/06/2020	Solid	L4292-04	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B13-100620-0-10	10/06/2020	9045D
10/06/2020	Solid	L4292-10	Corrosivity	Cool 4 deg C	ENTE01	O21	PAV-B13-100620-DP	10/06/2020	9045D
10/07/2020	Solid	L4311-04	Corrosivity	Cool 4 deg C	ENTE01	O31	PAV-B10-100720-0-10	10/07/2020	9045D
10/07/2020	Solid	L4311-06	Corrosivity	Cool 4 deg C	ENTE01	O31	PAV-B10-100720-10-17	10/07/2020	9045D

Date/Time 10-07-2020 16:25
 Received by: JS (wetchem)
 Relinquished by: JS (wetchem)

Date/Time 10-07-2020 18:19
 Received by: JS (wetchem)
 Relinquished by: JS (wetchem)