


LB111371

INSTRUMENT ID: CY1

7471B

Linear



A= 0.0000e+000

B= 1.9348e-004

C= -3.4975e-002

Rho= 0.9997611

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	76	0.000	76				
0.2	0.200	0.224	0.024	1340	0.0 %	1340				
2.5	2.500	2.395	-0.105	12558	0.0 %	12558				
5.0	5.000	5.136	0.136	26726	0.0 %	26726				
7.5	7.500	7.545	0.045	39179	0.0 %	39179				
10.0	10.000	9.920	-0.080	51452	0.0 %	51452				

LB111371 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 08 Oct 2020 10:49:10				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	76	-	0.0000	7471B	PPB	08 Oct 2020 10:53:42	S Std
0.2 - 1	50.2	1340	-	0.2000	7471B	PPB	08 Oct 2020 10:56:00	S Std
2.5 - 1	52.5	12558	-	2.5000	7471B	PPB	08 Oct 2020 10:58:17	S Std
5.0 - 1	55	26726	-	5.0000	7471B	PPB	08 Oct 2020 11:00:34	S Std
7.5 - 1	57.5	39179	-	7.5000	7471B	PPB	08 Oct 2020 11:02:51	S Std
10.0 - 1	510	51452	-	10.0000	7471B	PPB	08 Oct 2020 11:10:17	S Std
ICV47 - 1	ICV47	19992	3.8331	-	7471B	PPB	08 Oct 2020 11:12:55	U SMPL
ICB47 - 1	ICB47	327	0.0283	-	7471B	PPB	08 Oct 2020 11:15:09	U SMPL
CCV80 - 1	CCV80	25619	4.9218	-	7471B	PPB	08 Oct 2020 11:17:25	U SMPL
CCB80 - 1	CCB80	337	0.0302	-	7471B	PPB	08 Oct 2020 11:19:40	U SMPL
CRA - 1	CRA	1412	0.2382	-	7471B	PPB	08 Oct 2020 11:21:56	U SMPL
HighStd - 1	HighStd	50793	9.7924	-	7471B	PPB	08 Oct 2020 11:24:11	U SMPL
ChkStd - 1	ChkStd	36955	7.1151	-	7471B	PPB	08 Oct 2020 11:29:07	U SMPL
PB132194BL - 1	PBS	288	0.0207	-	7471B	PPB	08 Oct 2020 11:31:22	U SMPL
PB132194BS - 1	LCSS	17959	3.4397	-	7471B	PPB	08 Oct 2020 11:33:43	U SMPL
L4221-06 - 1	OK-01-100620	3335	0.6103	-	7471B	PPB	08 Oct 2020 11:36:00	U SMPL
L4224-01 - 1	OR-02-100620	2551	0.4586	-	7471B	PPB	08 Oct 2020 11:38:17	U SMPL
L4225-01 - 1	SU-04-100620	1694	0.2928	-	7471B	PPB	08 Oct 2020 11:45:29	U SMPL
L4225-01DUP - 1	SU-04-100620DUP	1686	0.2912	-	7471B	PPB	08 Oct 2020 11:47:46	U SMPL
L4225-01MS - 1	SU-04-100620MS	20999	4.0279	-	7471B	PPB	08 Oct 2020 11:50:01	U SMPL
CCV81 - 1	CCV81	25403	4.8800	-	7471B	PPB	08 Oct 2020 11:52:16	U SMPL
CCB81 - 1	CCB81	338	0.0304	-	7471B	PPB	08 Oct 2020 11:54:31	U SMPL
L4225-01MSD - 1	SU-04-100620MSD	21845	4.1916	-	7471B	PPB	08 Oct 2020 11:56:47	U SMPL
L4279-03 - 1	S-2	1511	0.2574	-	7471B	PPB	08 Oct 2020 11:59:03	U SMPL
L4281-03 - 1	49-93	745	0.1092	-	7471B	PPB	08 Oct 2020 12:01:19	U SMPL
L4282-03 - 1	68-89	858	0.1310	-	7471B	PPB	08 Oct 2020 12:03:35	U SMPL
L4291-03 - 1	PAV-B4-100620-0-10	19778	3.7917	-	7471B	PPB	08 Oct 2020 12:05:50	U SMPL
L4291-05 - 1	PAV-B4-100620-10-20	6530	1.2284	-	7471B	PPB	08 Oct 2020 12:08:07	U SMPL
L4291-11 - 1	PAV-B4-100620-DP	17756	3.4004	-	7471B	PPB	08 Oct 2020 12:10:24	U SMPL
L4292-03 - 1	PAV-B13-100620-0-10	29308	5.6355	-	7471B	PPB	08 Oct 2020 12:12:40	U SMPL
L4292-09 - 1	PAV-B13-100620-DP	26264	5.0466	-	7471B	PPB	08 Oct 2020 12:14:57	U SMPL
L4298-02 - 1	RBR-559	6392	1.2017	-	7471B	PPB	08 Oct 2020 12:17:14	U SMPL
CCV82 - 1	CCV82	26037	5.0026	-	7471B	PPB	08 Oct 2020 12:19:30	U SMPL
CCB82 - 1	CCB82	362	0.0351	-	7471B	PPB	08 Oct 2020 12:21:45	U SMPL
L4309-01 - 1	OWLS-HEAD-BH-02-A	9877	1.8760	-	7471B	PPB	08 Oct 2020 12:24:01	U SMPL
L4309-02 - 1	OWLS-HEAD-BH-02-B	7444	1.4053	-	7471B	PPB	08 Oct 2020 12:26:17	U SMPL
L4309-03 - 1	OWLS-HEAD-BH-04-A	2983	0.5422	-	7471B	PPB	08 Oct 2020 12:28:32	U SMPL
L4309-04 - 1	OWLS-HEAD-BH-04-B	4336	0.8040	-	7471B	PPB	08 Oct 2020 12:30:48	U SMPL
L4309-05 - 1	OWLS-HEAD-BH-05-A	96077	18.5540	-	7471B	PPB	08 Oct 2020 12:33:03	U SMPL
L4309-06 - 1	OWLS-HEAD-BH-05-B	125954	24.3345	-	7471B	PPB	08 Oct 2020 12:35:20	U SMPL
PB132195BL - 1	PBS	127	-0.0104	-	7471B	PPB	08 Oct 2020 12:37:39	U SMPL
PB132195BS - 1	LCSS	17491	3.3492	-	7471B	PPB	08 Oct 2020 12:39:59	U SMPL
L4294-01 - 1	TP-(0-3)-(3-5)-COMP	1629	0.2802	-	7471B	PPB	08 Oct 2020 12:42:16	U SMPL
L4297-01 - 1	RO-RBR-200031	1065	0.1711	-	7471B	PPB	08 Oct 2020 12:44:32	U SMPL
CCV83 - 1	CCV83	25124	4.8260	-	7471B	PPB	08 Oct 2020 12:46:47	U SMPL
CCB83 - 1	CCB83	45	-0.0263	-	7471B	PPB	08 Oct 2020 12:49:03	U SMPL
L4297-03 - 1	RO-RBR-200034	816	0.1229	-	7471B	PPB	08 Oct 2020 12:51:19	U SMPL
L4303-21 - 1	IDW-20201006	489	0.0596	-	7471B	PPB	08 Oct 2020 12:53:34	U SMPL
L4303-21DUP - 1	IDW-20201006DUP	402	0.0428	-	7471B	PPB	08 Oct 2020 12:55:50	U SMPL
L4303-21MS - 1	IDW-20201006MS	17191	3.2911	-	7471B	PPB	08 Oct 2020 12:58:06	U SMPL
L4303-21MSD - 1	IDW-20201006MSD	17933	3.4347	-	7471B	PPB	08 Oct 2020 13:00:22	U SMPL
L4311-03 - 1	PAV-B10-100720-0-10	11844	2.2566	-	7471B	PPB	08 Oct 2020 13:02:38	U SMPL
L4311-05 - 1	PAV-B10-100720-10-17	2964	0.5385	-	7471B	PPB	08 Oct 2020 13:04:55	U SMPL
CCV84 - 1	CCV84	25510	4.9007	-	7471B	PPB	08 Oct 2020 13:07:12	U SMPL
CCB84 - 1	CCB84	90	-0.0176	-	7471B	PPB	08 Oct 2020 13:09:29	U SMPL
L4225-01LX5 - 1		630	0.0869	-	7471B	PPB	08 Oct 2020 13:11:45	U SMPL
L4225-01A - 1		22071	4.2353	-	7471B	PPB	08 Oct 2020 13:14:00	U SMPL
L4303-21LX5 - 1		353	0.0333	-	7471B	PPB	08 Oct 2020 13:16:17	U SMPL
L4303-21A - 1		19335	3.7060	-	7471B	PPB	08 Oct 2020 13:18:33	U SMPL
L4309-05DLX5 - 1		18635	3.5705	-	7471B	PPB	08 Oct 2020 13:20:48	U SMPL
L4309-06DLX5 - 1		23827	4.5751	-	7471B	PPB	08 Oct 2020 13:23:04	U SMPL
CCV85 - 1	CCV85	25523	4.9032	-	7471B	PPB	08 Oct 2020 13:25:20	U SMPL
CCB85 - 1	CCB85	196	0.0029	-	7471B	PPB	08 Oct 2020 13:27:36	U SMPL