



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111372

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/08/2020	09:20
2	CAL2	1	Water	NA	NA	20.1	7.00	10/08/2020	09:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/08/2020	09:22
4	ICV	1	Water	NA	NA	20.2	7.01	10/08/2020	09:24
5	CCV1	1	Water	NA	NA	20.1	2.01	10/08/2020	09:25
6	L4294-01	1	Solid	20.02	20	23.2	9.15	10/08/2020	09:28
7	L4294-01DUP	1	Solid	20.03	20	23.3	9.16	10/08/2020	09:29
8	CCV2	1	Water	NA	NA	20.1	12.02	10/08/2020	09:30

VB 111372

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4294 WorkList ID : 143089 Department : Wet-Chemistry Date : 10-08-2020 09:05:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/07/2020	Solid	L4294-01	pH	Cool 4 deg C	PSEG01	O41	TP-(0-3)-(3-5)-COMP	10/07/2020	9045D

Date/Time 10-08-2020 9:10
Received by: g / wet (chem)
Relinquished by: HA

Date/Time 10-08-2020 10:20
Received by: g / wet
Relinquished by: g / wet