



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111376

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/08/2020	09:45
2	CAL2	1	Water	NA	NA	20.2	7.00	10/08/2020	09:46
3	CAL3	1	Water	NA	NA	20.2	10.03	10/08/2020	09:47
4	ICV	1	Water	NA	NA	20.2	7.00	10/08/2020	09:48
5	CCV1	1	Water	NA	NA	20.1	2.01	10/08/2020	09:50
6	L4324-01	1	Water	NA	NA	19.4	7.73	10/08/2020	09:51
7	L4324-01DUP	1	Water	NA	NA	19.5	7.74	10/08/2020	09:52
8	CCV2	1	Water	NA	NA	20.3	12.02	10/08/2020	09:55

WORKLIST(Hardcopy Internal Chain)

VB 111376

WorkList Name : COROSIVITY L4303 WorkList ID : 143090 Department : Wet-Chemistry Date : 10-08-2020 09:06:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/06/2020	Water	L4303-19 -D	Corrosivity	Cool 4 deg C	TETR06	O31	IDW-20201006	10/06/2020	9040C
10/07/2020	Water	L4324-01	Corrosivity	Cool 4 deg C	SHAW22	O32	UST-CONTENT-LIQUID	10/07/2020	9040C

Date/Time 10-08-2020 9:40 (chem)
 Received by: SP f wet
 Relinquished by: CP smi

Date/Time 10-08-2020 11:00
 Received by: CP smi
 Relinquished by: SP f wet (chem)