

LB111380


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38647



A= 0.0000e+000

B= 2.6199e-004

C= -5.0291e-002

Rho= 0.9998359

Accept=Accepted

slope
+ intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.000	-0.000	191	0.000	191					10
0.05	0.050	0.055	0.005	400	0.0 %	400					-2
0.20	0.200	0.197	-0.003	943	0.0 %	943					0
2.50	2.500	2.511	0.011	9777	0.0 %	9777					1
5.00	5.000	5.062	0.062	19513	0.0 %	19513					-2
7.50	7.500	7.351	-0.149	28252	0.0 %	28252					1
10.0	10.000	10.075	0.075	38647	0.0 %	38647					

LB111380
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	191	-		0 ISM02.4	PPB	10/8/2020 14:27	Std
	0.05 S0.05	400	-	0.05	ISM02.4	PPB	10/8/2020 14:29	Std
	0.2 S0.2	943	-	0.2	ISM02.4	PPB	10/8/2020 14:31	Std
	2.5 S2.5	9777	-	2.5	ISM02.4	PPB	10/8/2020 14:33	Std
	5 S5.0	19513	-	5	ISM02.4	PPB	10/8/2020 14:36	Std
	7.5 S7.5	28252	-	7.5	ISM02.4	PPB	10/8/2020 14:38	Std
	10 S10.0	38647	-	10	ISM02.4	PPB	10/8/2020 14:43	Std
ICV	ICV	15539	4.0207 -		ISM02.4	PPB	10/8/2020 14:46	SMPL
ICB	ICB	155	-0.0097 -		ISM02.4	PPB	10/8/2020 14:51	SMPL
CCV086	CCV086	19260	4.9956 -		ISM02.4	PPB	10/8/2020 14:53	SMPL
CCB086	CCB086	146	-0.012 -		ISM02.4	PPB	10/8/2020 14:58	SMPL
PB132212BL	PBW212	143	-0.0128 -		ISM02.4	PPB	10/8/2020 15:00	SMPL
L4307-01	MEW0T0	192	0 -		ISM02.4	PPB	10/8/2020 15:03	SMPL
L4307-02	MEW0T2	89	-0.027 -		ISM02.4	PPB	10/8/2020 15:05	SMPL
L4307-03	MEW0T3	97	-0.0249 -		ISM02.4	PPB	10/8/2020 15:07	SMPL
L4307-04	MEW0T1	102	-0.0236 -		ISM02.4	PPB	10/8/2020 15:09	SMPL
L4307-05	MEW0T1D	93	-0.0259 -		ISM02.4	PPB	10/8/2020 15:12	SMPL
L4307-06	MEW0T1S	3983	0.9932 -		ISM02.4	PPB	10/8/2020 15:14	SMPL
CCV087	CCV087	19349	5.0189 -		ISM02.4	PPB	10/8/2020 15:16	SMPL
CCB087	CCB087	99	-0.0244 -		ISM02.4	PPB	10/8/2020 15:18	SMPL