


LB111885

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5483e-004

C= -1.1703e-002

Rho= 0.9998498

Accept=Accepted

Slope
x-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	197	0.000	197					=
0.05	0.050					400					19
0.20	0.200	0.237	0.037	975	0.0 %	975					-1
2.50	2.500	2.478	-0.022	9771	0.0 %	9771					-1
5.00	5.000	4.929	-0.071	19388	0.0 %	19388					-1
7.50	7.500	7.419	-0.081	29161	0.0 %	29161					-1
10.0	10.000	10.098	0.098	39674	0.0 %	39674					1

LB111385
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	197	-		0 ISM02.4	PPB	10/8/2020 15:40	Std
	0.2 S0.2	975	-		0.2 ISM02.4	PPB	10/8/2020 15:43	Std
	2.5 S2.5	9771	-		2.5 ISM02.4	PPB	10/8/2020 15:45	Std
	5 S5.0	19388	-		5 ISM02.4	PPB	10/8/2020 15:47	Std
	7.5 S7.5	29161	-		7.5 ISM02.4	PPB	10/8/2020 15:49	Std
	10 S10.0	39674	-		10 ISM02.4	PPB	10/8/2020 15:52	Std
ICV	ICV	15263	3.8777 -		ISM02.4	PPB	10/8/2020 15:55	SMPL
ICB	ICB	37	-0.0023 -		ISM02.4	PPB	10/8/2020 15:57	SMPL
CCV088	CCV088	18343	4.6626 -		ISM02.4	PPB	10/8/2020 15:59	SMPL
CCB088	CCB088	46	0 -		ISM02.4	PPB	10/8/2020 16:01	SMPL
PB132213BL	PBW213	166	0.0306 -		ISM02.4	PPB	10/8/2020 16:04	SMPL
L4207-15	MBEQ74	106	0.0153 -		ISM02.4	PPB	10/8/2020 16:06	SMPL
L4207-16	MBEQ77	97	0.013 -		ISM02.4	PPB	10/8/2020 16:08	SMPL
PB132214BL	PBW214	52	0.0015 -		ISM02.4	PPB	10/8/2020 16:10	SMPL
L4210-01	MJLWS7	47	0.0003 -		ISM02.4	PPB	10/8/2020 16:13	SMPL
L4210-02	MJLWS8	90	0.0112 -		ISM02.4	PPB	10/8/2020 16:15	SMPL
L4210-03	MJLWS8D	100	0.0138 -		ISM02.4	PPB	10/8/2020 16:17	SMPL
L4210-04	MJLWS8S	4072	1.026 -		ISM02.4	PPB	10/8/2020 16:19	SMPL
L4210-05	MJLWT0	84	0.0097 -		ISM02.4	PPB	10/8/2020 16:22	SMPL
L4210-06	MJLWS9	87	0.0105 -		ISM02.4	PPB	10/8/2020 16:24	SMPL
L4210-07	MJLWX4	378	0.0846 -		ISM02.4	PPB	10/8/2020 16:26	SMPL
CCV089	CCV089	18251	4.6392 -		ISM02.4	PPB	10/8/2020 16:28	SMPL
CCB089	CCB089	-5	-0.013 -		ISM02.4	PPB	10/8/2020 16:31	SMPL