



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111392

Slope : 98.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2629         |
| buffer solution pH 7 yellow    | W2627         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2632         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 10/08/2020 | 17:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 10/08/2020 | 17:01     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 10/08/2020 | 17:02     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 10/08/2020 | 17:05     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.01        | 10/08/2020 | 17:06     |
| 6   | L4331-01    | 1  | Water  | NA          | NA          | 21.8             | 8.45        | 10/08/2020 | 17:10     |
| 7   | L4331-01DUP | 1  | Water  | NA          | NA          | 21.9             | 8.46        | 10/08/2020 | 17:11     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.03       | 10/08/2020 | 17:12     |

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4332      WorkList ID : 143111      Department : Wet-Chemistry      Date : 10-08-2020 16:34:38

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 10/08/2020 | Water  | L4331-01 | pH   | Cool 4 deg C | TULL02   | O61              | SHELL-L15-L14   | 10/08/2020   | 9040C  |

Date/Time 10-08-2020 16:35  
 Received by: JD (wet-chem)  
 Relinquished by: KS (wet-chem)

Date/Time 10-08-2020 17:50  
 Received by: KS (wet-chem)  
 Relinquished by: JD (wet-chem)