



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111393

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/08/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	10/08/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/08/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	10/08/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	10/08/2020	17:06
6	L4314-01	1	Solid	20.03	20	21.7	7.26	10/08/2020	17:15
7	L4314-01DUP	1	Solid	20.04	20	21.8	7.27	10/08/2020	17:16
8	L4314-02	1	Solid	20.02	20	22.0	6.90	10/08/2020	17:19
9	L4314-03	1	Solid	20.03	20	21.7	8.71	10/08/2020	17:20
10	L4314-04	1	Solid	20.02	20	22.1	8.68	10/08/2020	17:21
11	CCV2	1	Water	NA	NA	20.2	12.02	10/08/2020	17:22

18111393

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH S L4314 WorkList ID: 143112 Department: Wet-Chemistry Date: 10-08-2020 16:35:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2020	Solid	L4314-01	pH	Cool 4 deg C	PSEG01	O51	31245	10/08/2020	9045D
10/08/2020	Solid	L4314-02	pH	Cool 4 deg C	PSEG01	O51	31246	10/08/2020	9045D
10/08/2020	Solid	L4314-03	pH	Cool 4 deg C	PSEG01	O51	31247	10/08/2020	9045D
10/08/2020	Solid	L4314-04	pH	Cool 4 deg C	PSEG01	O51	31248	10/08/2020	9045D

Date/Time 10-08-2020 16:40
 Received by: Jm (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 10-08-2020 18:10
 Received by: KS (wet-chem)
 Relinquished by: Jm (wet-chem)