



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111395

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/09/2020	09:05
2	L4296-02		1	8	100.00	>212.0	>212.0	10/09/2020	09:35
3	L4296-02DUP		1	8	100.00	>212.0	>212.0	10/09/2020	09:58
4	L4299-06		1	9	100.00	>212.0	>212.0	10/09/2020	10:20
5	L4300-02		1	9	100.00	>212.0	>212.0	10/09/2020	10:45
6	L4324-01		1	8	100.00	>212.0	>212.0	10/09/2020	11:10
7	L4331-01		1	8	100.00	>212.0	>212.0	10/09/2020	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111395

WorkList Name : fp10820

WorkList ID : 143080

Department : Wet-Chemistry

Date : 10-08-2020 08:56:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/20/2020	Water	L4296-02	Flash Point	Cool 4 deg C	PSEG01	O41	31281	10/07/2020	1010A
10/14/2020	Water	L4299-06	Flash Point	Cool 4 deg C	PSEG01	O41	86-96-COMP	10/07/2020	1010A
10/14/2020	Water	L4300-02	Flash Point	Cool 4 deg C	PSEG01	O42	223	10/07/2020	1010A
10/12/2020	Water	L4324-01	Flash Point	Cool 4 deg C	SHAW22	O32	UST-CONTENT-LIQUID	10/07/2020	1010A
10/09/2020	Water	L4331-01	Flash Point	Cool 4 deg C	TULL02	O61	SHELL-L15-L14	10/08/2020	1010A

Date/Time 10/9/2020 08:30
 Received by: KS (Webster)
 Relinquished by: KS (ext lab)

Date/Time 10/9/2020 11:45
 Received by: Fe (ext Lab)
 Relinquished by: KS (Webster)