

LB111397


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

32358



A= 0.0000e+000

B= 3.1029e-004

C= -2.1813e-002

Rho= 0.9997339

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	2	0.000	2				
0.2	0.200	0.161	-0.039	589	0.0 %	589				
2.5	2.500	2.534	0.034	8238	0.0 %	8238				
5.0	5.000	5.145	0.145	16651	0.0 %	16651				
7.5	7.500	7.362	-0.138	23797	0.0 %	23797				
10.0	10.000	10.019	0.019	32358	0.0 %	32358				

LB111397 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 09 Oct 2020 08:45:00				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	2	-	0.00007470A	PPB	09 Oct 2020 08:52:30	S	Std
0.2 - 1	50.2	589	-	0.20007470A	PPB	09 Oct 2020 08:54:47	S	Std
2.5 - 1	52.5	8238	-	2.50007470A	PPB	09 Oct 2020 08:57:04	S	Std
5.0 - 1	55	16651	-	5.00007470A	PPB	09 Oct 2020 08:59:21	S	Std
7.5 - 1	57.5	23797	-	7.50007470A	PPB	09 Oct 2020 09:01:37	S	Std
10.0 - 1	510	32358	-	10.00007470A	PPB	09 Oct 2020 09:03:53	S	Std
ICV48 - 1	ICV48	13152	4.0592	-7470A	PPB	09 Oct 2020 09:07:44	U	SMPL
ICB48 - 1	ICB48	-64	-0.0417	-7470A	PPB	09 Oct 2020 09:09:58	U	SMPL
CCV90 - 1	CCV90	16935	5.2330	-7470A	PPB	09 Oct 2020 09:12:13	U	SMPL
CCB90 - 1	CCB90	-69	-0.0432	-7470A	PPB	09 Oct 2020 09:14:28	U	SMPL
CRA - 1	CRA	817	0.2317	-7470A	PPB	09 Oct 2020 09:19:45	U	SMPL
HighStd - 1	HighStd	33489	10.3696	-7470A	PPB	09 Oct 2020 09:22:00	U	SMPL
ChkStd - 1	ChkStd	22719	7.0278	-7470A	PPB	09 Oct 2020 09:24:15	U	SMPL
PB132215BL - 1	PBW	-174	-0.0758	-7470A	PPB	09 Oct 2020 09:26:32	U	SMPL
PB132215BS - 1	LCSW	11205	3.4550	-7470A	PPB	09 Oct 2020 09:31:21	U	SMPL
L4291-04 - 1	PAV-B4-100620-0-10	-48	-0.0367	-7470A	PPB	09 Oct 2020 09:33:37	U	SMPL
L4291-04DUP - 1	PAV-B4-100620-0-10DUP	11	-0.0184	-7470A	PPB	09 Oct 2020 09:35:53	U	SMPL
L4291-04MS - 1	PAV-B4-100620-0-10MS	11270	3.4752	-7470A	PPB	09 Oct 2020 09:40:49	U	SMPL
L4291-04MSD - 1	PAV-B4-100620-0-10MSD	11703	3.6096	-7470A	PPB	09 Oct 2020 09:43:05	U	SMPL
L4291-06 - 1	PAV-B4-100620-10-20	-91	-0.0500	-7470A	PPB	09 Oct 2020 09:45:21	U	SMPL
CCV91 - 1	CCV91	15025	4.6404	-7470A	PPB	09 Oct 2020 09:47:36	U	SMPL
CCB91 - 1	CCB91	-92	-0.0504	-7470A	PPB	09 Oct 2020 09:49:52	U	SMPL
L4291-12 - 1	PAV-B4-100620-DP	-14	-0.0262	-7470A	PPB	09 Oct 2020 09:52:07	U	SMPL
L4292-04 - 1	PAV-B13-100620-0-10	0	-0.0218	-7470A	PPB	09 Oct 2020 09:54:23	U	SMPL
L4292-10 - 1	PAV-B13-100620-DP	19	-0.0159	-7470A	PPB	09 Oct 2020 09:56:38	U	SMPL
L4294-02 - 1	TP-(0-3)-(3-5)-COMP	6	-0.0200	-7470A	PPB	09 Oct 2020 09:58:54	U	SMPL
L4295-01 - 1	DRUM-1	26	-0.0137	-7470A	PPB	09 Oct 2020 10:01:09	U	SMPL
L4295-02 - 1	10-07A	22	-0.0150	-7470A	PPB	09 Oct 2020 10:03:26	U	SMPL
L4295-03 - 1	10-07B	67	-0.0010	-7470A	PPB	09 Oct 2020 10:07:07	U	SMPL
L4295-04 - 1	10-07C	33	-0.0116	-7470A	PPB	09 Oct 2020 10:09:23	U	SMPL
L4300-01 - 1	231-232-COMP	45	-0.0078	-7470A	PPB	09 Oct 2020 10:11:41	U	SMPL
L4311-04 - 1	PAV-B10-100720-0-10	44	-0.0082	-7470A	PPB	09 Oct 2020 10:13:57	U	SMPL
CCV92 - 1	CCV92	17678	5.4636	-7470A	PPB	09 Oct 2020 10:19:21	U	SMPL
CCB92 - 1	CCB92	-70	-0.0435	-7470A	PPB	09 Oct 2020 10:21:36	U	SMPL
L4311-06 - 1	PAV-B10-100720-10-17	23	-0.0147	-7470A	PPB	09 Oct 2020 10:23:51	U	SMPL
PB132216BL - 1	PBW	27	-0.0134	-7470A	PPB	09 Oct 2020 10:26:07	U	SMPL
PB132216BS - 1	LCSW	12882	3.9754	-7470A	PPB	09 Oct 2020 10:28:24	U	SMPL
L4296-02 - 1	31281	73	0.0008	-7470A	PPB	09 Oct 2020 10:30:39	U	SMPL
L4299-06 - 1	86-96-COMP	2	-0.0212	-7470A	PPB	09 Oct 2020 10:32:55	U	SMPL
L4299-06DUP - 1	86-96-COMPDUP	12	-0.0181	-7470A	PPB	09 Oct 2020 10:35:11	U	SMPL
L4299-06MSD - 1	86-96-COMPMSD	11355	3.5016	-7470A	PPB	09 Oct 2020 10:42:20	U	SMPL
L4300-02 - 1	223	-35	-0.0327	-7470A	PPB	09 Oct 2020 10:44:38	U	SMPL
L4299-06MS - 1	86-96-COMPMS	10935	3.3713	-7470A	PPB	09 Oct 2020 10:47:12	U	SMPL
CCV93 - 1	CCV93	16132	4.9839	-7470A	PPB	09 Oct 2020 10:49:55	U	SMPL
CCB93 - 1	CCB93	-70	-0.0435	-7470A	PPB	09 Oct 2020 10:52:10	U	SMPL
PB132215TB - 1	PB132188TB	-10	-0.0249	-7470A	PPB	09 Oct 2020 10:54:26	U	SMPL
PB132216TB - 1	PB132192TB	34	-0.0113	-7470A	PPB	09 Oct 2020 10:56:42	U	SMPL
L4291-04LX5 - 1		83	0.0039	-7470A	PPB	09 Oct 2020 10:58:58	U	SMPL
L4291-04A - 1		14513	4.4815	-7470A	PPB	09 Oct 2020 11:01:13	U	SMPL
L4299-06LX5 - 1		-18	-0.0274	-7470A	PPB	09 Oct 2020 11:03:29	U	SMPL
L4299-06A - 1		14694	4.5377	-7470A	PPB	09 Oct 2020 11:08:22	U	SMPL
CCV94 - 1	CCV94	16785	5.1865	-7470A	PPB	09 Oct 2020 11:10:37	U	SMPL
CCB94 - 1	CCB94	-13	-0.0258	-7470A	PPB	09 Oct 2020 11:12:54	U	SMPL