



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111405

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/09/2020	13:35
2	CAL2	1	Water	NA	NA	20.1	7.00	10/09/2020	13:36
3	CAL3	1	Water	NA	NA	20.1	10.03	10/09/2020	13:37
4	ICV	1	Water	NA	NA	20.1	7.00	10/09/2020	13:40
5	CCV1	1	Water	NA	NA	20.2	2.01	10/09/2020	13:42
6	L4355-01	1	Water	NA	NA	19.3	6.21	10/09/2020	13:43
7	L4355-01DUP	1	Water	NA	NA	19.4	6.22	10/09/2020	13:44
8	CCV2	1	Water	NA	NA	20.1	12.02	10/09/2020	13:45

1B111405

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4356

Department : Wet-Chemistry

WorkList ID : 143154

Date : 10-09-2020 13:23:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/06/2020	Water	L4355-01	pH	Cool 4 deg C	STER14	O51	JS-SYSTEM	10/06/2020	SM 4500-H B

Date/Time 10-09-2020 13:30
 Received by: JA (wet-chem)
 Relinquished by: JA (wet-chem)

Date/Time 10-09-2020 14:30
 Received by: JA (wet-chem)
 Relinquished by: JA (wet-chem)