



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111406

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/09/2020	13:35
2	CAL2	1	Water	NA	NA	20.1	7.00	10/09/2020	13:36
3	CAL3	1	Water	NA	NA	20.1	10.03	10/09/2020	13:37
4	ICV	1	Water	NA	NA	20.1	7.00	10/09/2020	13:40
5	CCV1	1	Water	NA	NA	20.2	2.01	10/09/2020	13:42
6	L4333-01	1	Solid	20.03	20	19.9	8.04	10/09/2020	13:48
7	L4333-01DUP	1	Solid	20.04	20	19.8	8.05	10/09/2020	13:49
8	L4333-05	1	Solid	20.02	20	19.9	11.47	10/09/2020	13:51
9	L4333-08	1	Solid	20.03	20	19.9	10.94	10/09/2020	13:55
10	L4333-11	1	Solid	20.01	20	19.9	11.82	10/09/2020	14:00
11	L4333-14	1	Solid	20.03	20	20.0	11.44	10/09/2020	14:02
12	CCV2	1	Water	NA	NA	20.2	12.02	10/09/2020	14:05

LB 111406

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4333 WorkList ID : 143153 Department : Wet-Chemistry Date : 10-09-2020 13:23:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2020	Solid	L4333-01	pH	Cool 4 deg C	ENTA05	O51	OH-WC-J6-3-6	10/08/2020	9045D
10/08/2020	Solid	L4333-05	pH	Cool 4 deg C	ENTA05	O51	OH-WC-C-J4	10/08/2020	9045D
10/08/2020	Solid	L4333-08	pH	Cool 4 deg C	ENTA05	O51	OH-WC-C-G2	10/08/2020	9045D
10/08/2020	Solid	L4333-11	pH	Cool 4 deg C	ENTA05	O51	OH-WC-C-C1	10/08/2020	9045D
10/08/2020	Solid	L4333-14	pH	Cool 4 deg C	ENTA05	O51	OH-WC-C-D4	10/08/2020	9045D

Date/Time 10-09-2020 13:20
 Received by: SP (Wet-Chem)
 Relinquished by: AM (Wet-Chem)

Date/Time 10-09-2020 15:00
 Received by: AM (Wet-Chem)
 Relinquished by: JA (Wet-Chem)