

LB111408


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37526



A= 0.0000e+000

B= 2.6609e-004

C= 5.8763e-002

Rho= 0.9998197

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.061	0.061	7	0.000	7				
0.2	0.200	0.238	0.038	675	0.0 %	675				
2.5	2.500	2.448	-0.052	8978	0.0 %	8978				
5.0	5.000	4.867	-0.133	18068	0.0 %	18068				
7.5	7.500	7.543	0.043	28125	0.0 %	28125				
10.0	10.000	10.044	0.044	37526	0.0 %	37526				

LB111408 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 09 Oct 2020 15:16:51							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	09 Oct 2020 15:17:44	S	Std
0.2 - 1	50.2	675	-	0.2000	7470A	PPB	09 Oct 2020 15:20:01	S	Std
2.5 - 1	52.5	8978	-	2.5000	7470A	PPB	09 Oct 2020 15:22:18	S	Std
5.0 - 1	55	18068	-	5.0000	7470A	PPB	09 Oct 2020 15:24:35	S	Std
7.5 - 1	57.5	28125	-	7.5000	7470A	PPB	09 Oct 2020 15:26:52	S	Std
10.0 - 1	510	37526	-	10.0000	7470A	PPB	09 Oct 2020 15:29:08	S	Std
ICV50 - 1	ICV50	14888	4.0203	-	7470A	PPB	09 Oct 2020 15:34:23	U	SMPL
ICB50 - 1	ICB50	-107	0.0303	-	7470A	PPB	09 Oct 2020 15:36:38	U	SMPL
CCV01 - 1	CCV01	18006	4.8500	-	7470A	PPB	09 Oct 2020 16:14:06	U	SMPL
CCB01 - 1	CCB01	-41	0.0479	-	7470A	PPB	09 Oct 2020 16:16:21	U	SMPL
CRA - 1	CRA	632	0.2269	-	7470A	PPB	09 Oct 2020 16:18:36	U	SMPL
HighStd - 1	HighStd	37068	9.9223	-	7470A	PPB	09 Oct 2020 16:20:51	U	SMPL
ChkStd - 1	ChkStd	25002	6.7116	-	7470A	PPB	09 Oct 2020 16:23:06	U	SMPL
PB132250BL - 1	PBW	-260	-0.0104	-	7470A	PPB	09 Oct 2020 16:25:23	U	SMPL
PB132250BS - 1	LCSW	13099	3.5443	-	7470A	PPB	09 Oct 2020 16:27:40	U	SMPL
L4314-05 - 1	DRUM-1-4	-77	0.0383	-	7470A	PPB	09 Oct 2020 16:29:56	U	SMPL
L4314-05DUP - 1	DRUM-1-4DUP	20	0.0641	-	7470A	PPB	09 Oct 2020 16:32:13	U	SMPL
L4314-05MS - 1	DRUM-1-4MS	13151	3.5581	-	7470A	PPB	09 Oct 2020 16:34:30	U	SMPL
L4314-05MSD - 1	DRUM-1-4MSD	13197	3.5704	-	7470A	PPB	09 Oct 2020 16:36:47	U	SMPL
L4316-02 - 1	11929	-24	0.0524	-	7470A	PPB	09 Oct 2020 16:39:02	U	SMPL
CCV02 - 1	CCV02	18989	5.1116	-	7470A	PPB	09 Oct 2020 16:41:17	U	SMPL
CCB02 - 1	CCB02	-145	0.0202	-	7470A	PPB	09 Oct 2020 16:43:32	U	SMPL
L4316-04 - 1	26414	-27	0.0516	-	7470A	PPB	09 Oct 2020 16:45:47	U	SMPL
L4317-02 - 1	11991	5	0.0601	-	7470A	PPB	09 Oct 2020 16:48:03	U	SMPL
L4317-04 - 1	11935	-10	0.0561	-	7470A	PPB	09 Oct 2020 16:50:18	U	SMPL
L4318-02 - 1	214	25	0.0654	-	7470A	PPB	09 Oct 2020 16:52:34	U	SMPL
L4319-02 - 1	11995	16	0.0630	-	7470A	PPB	09 Oct 2020 16:54:49	U	SMPL
L4319-04 - 1	12018	29	0.0665	-	7470A	PPB	09 Oct 2020 16:57:06	U	SMPL
L4332-02 - 1	SHELL-L15-L14	294	0.1370	-	7470A	PPB	09 Oct 2020 16:59:22	U	SMPL
PB132251BL - 1	PBW	-31	0.0505	-	7470A	PPB	09 Oct 2020 17:01:39	U	SMPL
PB132251BS - 1	LCSW	13082	3.5398	-	7470A	PPB	09 Oct 2020 17:03:56	U	SMPL
L4181-14 - 1	40371	82	0.0806	-	7470A	PPB	09 Oct 2020 17:09:45	U	SMPL
CCV03 - 1	CCV03	19338	5.2045	-	7470A	PPB	09 Oct 2020 17:11:59	U	SMPL
CCB03 - 1	CCB03	-35	0.0494	-	7470A	PPB	09 Oct 2020 17:14:15	U	SMPL
L4181-14DUP - 1	40371DUP	63	0.0755	-	7470A	PPB	09 Oct 2020 17:16:30	U	SMPL
L4181-14MS - 1	40371MS	13705	3.7056	-	7470A	PPB	09 Oct 2020 17:18:46	U	SMPL
L4181-14MSD - 1	40371MSD	13964	3.7745	-	7470A	PPB	09 Oct 2020 17:21:01	U	SMPL
PB132252BL - 1	PBW	-63	0.0420	-	7470A	PPB	09 Oct 2020 17:23:17	U	SMPL
PB132252BS - 1	LCSW	12985	3.5140	-	7470A	PPB	09 Oct 2020 17:25:33	U	SMPL
L4291-13 - 1	FIELD-BLANK	-160	0.0162	-	7470A	PPB	09 Oct 2020 17:27:49	U	SMPL
L4292-11 - 1	FIELD-BLANK	-64	0.0417	-	7470A	PPB	09 Oct 2020 17:33:11	U	SMPL
L4324-01 - 1	UST-CONTENT-LIQUID	558	0.2072	-	7470A	PPB	09 Oct 2020 17:35:28	U	SMPL
L4324-01DUP - 1	UST-CONTENT-LIQUIDDUP	509	0.1942	-	7470A	PPB	09 Oct 2020 17:40:18	U	SMPL
L4324-01MS - 1	UST-CONTENT-LIQUIDMS	13098	3.5440	-	7470A	PPB	09 Oct 2020 17:45:08	U	SMPL
CCV04 - 1	CCV04	19113	5.1446	-	7470A	PPB	09 Oct 2020 17:47:24	U	SMPL
CCB04 - 1	CCB04	-70	0.0401	-	7470A	PPB	09 Oct 2020 17:49:39	U	SMPL
L4324-01MSD - 1	UST-CONTENT-LIQUIDMSD	13758	3.7197	-	7470A	PPB	09 Oct 2020 17:51:55	U	SMPL
PB132250TB - 1	PB132228TB	-88	0.0353	-	7470A	PPB	09 Oct 2020 17:54:10	U	SMPL
PB132251TB - 1	PB132231TB	-25	0.0521	-	7470A	PPB	09 Oct 2020 17:56:26	U	SMPL
L4314-05LX5 - 1		99	0.0851	-	7470A	PPB	09 Oct 2020 17:58:42	U	SMPL
L4314-05A - 1		14733	3.9791	-	7470A	PPB	09 Oct 2020 18:00:57	U	SMPL
L4181-14LX5 - 1		-6	0.0572	-	7470A	PPB	09 Oct 2020 18:03:13	U	SMPL
L4181-14A - 1		14850	4.0102	-	7470A	PPB	09 Oct 2020 18:05:30	U	SMPL
L4324-01LX5 - 1		10	0.0614	-	7470A	PPB	09 Oct 2020 18:07:46	U	SMPL
L4324-01A - 1		14609	3.9461	-	7470A	PPB	09 Oct 2020 18:10:03	U	SMPL
CCV05 - 1	CCV05	18806	5.0629	-	7470A	PPB	09 Oct 2020 18:12:19	U	SMPL
CCB05 - 1	CCB05	-83	0.0367	-	7470A	PPB	09 Oct 2020 18:14:35	U	SMPL