



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111413

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/09/2020	16:50
2	CAL2	1	Water	NA	NA	20.1	7.00	10/09/2020	16:51
3	CAL3	1	Water	NA	NA	20.1	10.03	10/09/2020	16:52
4	ICV	1	Water	NA	NA	20.1	7.00	10/09/2020	16:55
5	CCV1	1	Water	NA	NA	20.2	2.03	10/09/2020	16:56
6	L4225-09	1	Solid	20.03	20	21.4	7.50	10/09/2020	17:00
7	L4225-09DUP	1	Solid	20.04	20	21.5	7.51	10/09/2020	17:01
8	L4333-02	1	Solid	20.03	20	19.9	8.04	10/09/2020	17:02
9	L4333-06	1	Solid	20.03	20	19.9	11.47	10/09/2020	17:05
10	L4333-09	1	Solid	20.02	20	19.4	10.94	10/09/2020	17:06
11	L4333-12	1	Solid	20.03	20	19.9	11.82	10/09/2020	17:10
12	L4333-15	1	Solid	20.04	20	20.0	11.44	10/09/2020	17:11
13	L4340-02	1	Solid	20.02	20	20.3	9.20	10/09/2020	17:13
14	L4340-04	1	Solid	20.04	20	21.5	8.97	10/09/2020	17:15
15	L4356-01	1	Solid	20.02	20	19.8	6.75	10/09/2020	17:19
16	CCV2	1	Water	NA	NA	20.2	12.03	10/09/2020	17:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4333 WorkList ID : 143149 Department : Wet-Chemistry Date : 10-09-2020 11:13:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2020	Solid	L4225-09	Corrosivity	Cool 4 deg C	PSEG05		SU-01-100920	10/09/2020	9045D
10/08/2020	Solid	L4333-02	Corrosivity	Cool 4 deg C	ENTA05	O51	OH-WC-J6-3-6	10/08/2020	9045D
10/08/2020	Solid	L4333-06	Corrosivity	Cool 4 deg C	ENTA05	O51	OH-WC-C-J4	10/08/2020	9045D
10/08/2020	Solid	L4333-09	Corrosivity	Cool 4 deg C	ENTA05	O51	OH-WC-C-G2	10/08/2020	9045D
10/08/2020	Solid	L4333-12	Corrosivity	Cool 4 deg C	ENTA05	O51	OH-WC-C-C1	10/08/2020	9045D
10/08/2020	Solid	L4333-15	Corrosivity	Cool 4 deg C	ENTA05	O51	OH-WC-C-D4	10/08/2020	9045D
10/09/2020	Solid	L4340-02	Corrosivity	Cool 4 deg C	PSEG01	O52	A-2332-MH11	10/09/2020	9045D
10/09/2020	Solid	L4340-04	Corrosivity	Cool 4 deg C	PSEG01	O52	A-2332-MH12	10/09/2020	9045D
10/08/2020	Solid	L4356-01	Corrosivity	Cool 4 deg C	CLEA11	O51	RECLAMATION-TANK	10/08/2020	9045D

Date/Time 10-09-2020 16:00
 Received by: JS (Wet-Chem)
 Relinquished by: JS (Wet-Chem)

Date/Time 10-09-2020 18:30
 Received by: JS (Wet-Chem)
 Relinquished by: JS (Wet-Chem)