


LB111498

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8783e-004

C= 2.1317e-002

Rho= 0.9991526

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	47	0.000	47				
0.2	0.200	0.228	0.028	1099	0.0 %	1099				
2.5	2.500	2.469	-0.031	13031	0.0 %	13031				
5.0	5.000	5.114	0.114	27111	0.0 %	27111				
7.5	7.500	7.192	-0.308	38177	0.0 %	38177				
10.0	10.000	10.168	0.168	54018	0.0 %	54018				

LB111428 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Oct 2020 09:27:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471B	PPB		12 Oct 2020 09:51:37	S	Std
0.2 - 1	50.2	1099	-	0.20007471B	PPB		12 Oct 2020 09:53:55	S	Std
2.5 - 1	52.5	13031	-	2.50007471B	PPB		12 Oct 2020 09:56:12	S	Std
5.0 - 1	55	27111	-	5.00007471B	PPB		12 Oct 2020 09:58:29	S	Std
7.5 - 1	57.5	38177	-	7.50007471B	PPB		12 Oct 2020 10:03:23	S	Std
10.0 - 1	510	54018	-	10.00007471B	PPB		12 Oct 2020 10:08:15	S	Std
ICV51 - 1	ICV51	21455	4.0512	-7471B	PPB		12 Oct 2020 10:14:22	U	SMPL
ICB51 - 1	ICB51	-74	0.0074	-7471B	PPB		12 Oct 2020 10:16:37	U	SMPL
CCV06 - 1	CCV06	26919	5.0775	-7471B	PPB		12 Oct 2020 10:18:53	U	SMPL
CCB06 - 1	CCB06	-81	0.0061	-7471B	PPB		12 Oct 2020 10:21:08	U	SMPL
CRA - 1	CRA	1057	0.2199	-7471B	PPB		12 Oct 2020 10:23:24	U	SMPL
HighStd - 1	HighStd	53918	10.1487	-7471B	PPB		12 Oct 2020 10:25:39	U	SMPL
ChkStd - 1	ChkStd	36443	6.8664	-7471B	PPB		12 Oct 2020 10:27:54	U	SMPL
PB132275BL - 1	PBS	-140	-0.0050	-7471B	PPB		12 Oct 2020 10:30:12	U	SMPL
PB132275BS - 1	LCSS	19056	3.6006	-7471B	PPB		12 Oct 2020 10:35:02	U	SMPL
L4225-05 - 1	SU-01-100920	2197	0.4340	-7471B	PPB		12 Oct 2020 10:37:18	U	SMPL
L4315-02 - 1	346	1581	0.3183	-7471B	PPB		12 Oct 2020 10:39:34	U	SMPL
L4316-01 - 1	11929	1227	0.2518	-7471B	PPB		12 Oct 2020 10:41:51	U	SMPL
L4316-03 - 1	26414	9885	1.8780	-7471B	PPB		12 Oct 2020 10:44:08	U	SMPL
L4317-01 - 1	11991	586	0.1314	-7471B	PPB		12 Oct 2020 10:46:23	U	SMPL
CCV07 - 1	CCV07	25290	4.7715	-7471B	PPB		12 Oct 2020 10:48:39	U	SMPL
CCB07 - 1	CCB07	-67	0.0087	-7471B	PPB		12 Oct 2020 10:50:54	U	SMPL
L4317-03 - 1	11935	1354	0.2756	-7471B	PPB		12 Oct 2020 10:53:10	U	SMPL
L4318-01 - 1	214	1677	0.3363	-7471B	PPB		12 Oct 2020 10:55:25	U	SMPL
L4319-01 - 1	11995	1745	0.3491	-7471B	PPB		12 Oct 2020 10:57:40	U	SMPL
L4319-01DUP - 1	11995DUP	1796	0.3587	-7471B	PPB		12 Oct 2020 10:59:56	U	SMPL
L4319-01MS - 1	11995MS	20290	3.8324	-7471B	PPB		12 Oct 2020 11:04:50	U	SMPL
L4319-01MSD - 1	11995MSD	20466	3.8654	-7471B	PPB		12 Oct 2020 11:07:06	U	SMPL
L4319-03 - 1	12018	1643	0.3299	-7471B	PPB		12 Oct 2020 11:09:22	U	SMPL
L4329-01 - 1	S-1	1672	0.3354	-7471B	PPB		12 Oct 2020 11:11:40	U	SMPL
L4332-01 - 1	SHELL-L15-L14	25804	4.8681	-7471B	PPB		12 Oct 2020 11:13:56	U	SMPL
L4337-02 - 1	TP-4	2528	0.4962	-7471B	PPB		12 Oct 2020 11:16:12	U	SMPL
CCV08 - 1	CCV08	26052	4.9147	-7471B	PPB		12 Oct 2020 11:18:30	U	SMPL
CCB08 - 1	CCB08	-54	0.0112	-7471B	PPB		12 Oct 2020 11:20:45	U	SMPL
L4340-01 - 1	A-2332-MH11	897	0.1898	-7471B	PPB		12 Oct 2020 11:23:01	U	SMPL
L4340-03 - 1	A-2332-MH12	9215	1.7522	-7471B	PPB		12 Oct 2020 11:25:17	U	SMPL
L4341-02 - 1	STOCK-PILE	1572	0.3166	-7471B	PPB		12 Oct 2020 11:27:32	U	SMPL
L4342-06 - 1	BACKFILL-OFFSITE-1-SO-0001	389	0.0944	-7471B	PPB		12 Oct 2020 11:29:48	U	SMPL
L4342-07 - 1	BACKFILL-OFFSITE-2-SO-0002	392	0.0949	-7471B	PPB		12 Oct 2020 11:32:04	U	SMPL
L4342-08 - 1	BACKFILL-OFFSITE-3-SO-0003	351	0.0872	-7471B	PPB		12 Oct 2020 11:34:20	U	SMPL
L4342-09 - 1	BACKFILL-OFFSITE-4-SO-0004	387	0.0940	-7471B	PPB		12 Oct 2020 11:36:36	U	SMPL
L4342-10 - 1	BACKFILL-OFFSITE-5-SO-0005	382	0.0931	-7471B	PPB		12 Oct 2020 11:38:53	U	SMPL
L4319-01LX5 - 1		525	0.1199	-7471B	PPB		12 Oct 2020 11:41:11	U	SMPL
L4319-01A - 1		21812	4.1183	-7471B	PPB		12 Oct 2020 11:43:26	U	SMPL
CCV09 - 1	CCV09	25534	4.8174	-7471B	PPB		12 Oct 2020 11:45:42	U	SMPL
CCB09 - 1	CCB09	-55	0.0110	-7471B	PPB		12 Oct 2020 11:47:58	U	SMPL