



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111438

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/12/2020	14:11
2	CAL2	1	Water	NA	NA	20.1	7.00	10/12/2020	14:12
3	CAL3	1	Water	NA	NA	20.1	10.03	10/12/2020	14:15
4	ICV	1	Water	NA	NA	20.1	7.00	10/12/2020	14:16
5	CCV1	1	Water	NA	NA	20.2	2.02	10/12/2020	14:19
6	L4378-01	1	Water	NA	NA	18.0	5.20	10/12/2020	14:20
7	L4378-01DUP	1	Water	NA	NA	18.1	5.21	10/12/2020	14:21
8	CCV2	1	Water	NA	NA	20.3	12.02	10/12/2020	14:25

18 111438

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4378 WorkList ID : 143209 Department : Wet-Chemistry Date : 10-12-2020 14:08:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2020	Water	L4378-01	pH	Cool 4 deg C	STER11	O62	POOL-WATER	10/09/2020	SM 4500-H B

Date/Time 10-12-2020 14:10
Received by: JN (wet-chem)
Relinquished by: JN (wet-chem)

Date/Time 10-12-2020 15:10
Received by: JN (wet-chem)
Relinquished by: JN (wet-chem)