



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111443

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP87441
Turbidity Calibration std, 0NTU	WP87440
Turbidity Calibration std, 20NTU	WP87436
Turbidity Calibration std, 40NTU	WP87435
Turbidity Calibration std, 1NTU	WP87439
Turbidity Calibration std, 5NTU	WP87438
Turbidity LOQ std, 1.0NTU	WP87442
Turbidity Calibration - CCV std, 10 NTU	WP87437
10 NTU Standard 500 ml	W2489
TURBIDITY LOD STD, 0.5NTU	WP87441
Turbidity LOQ std, 1.0NTU	WP87442

Intercept: 0.1758 Slope: 0.981 Regression: 0.9998						
Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.251	10/12/2020	11:40
2	CAL2	1	1	1.220	10/12/2020	11:42
3	CAL3	5	1	5.320	10/12/2020	11:44
4	CAL4	10	1	9.920	10/12/2020	11:47
5	CAL5	20	1	19.200	10/12/2020	11:49
6	CAL6	40	1	39.700	10/12/2020	11:52



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111443

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.219	10/12/2020	11:54
2	ICB		1	0.282	0.109	10/12/2020	11:56
3	CCV1	10	1	9.790	9.801	10/12/2020	11:58
4	CCB1		1	0.289	0.116	10/12/2020	12:01
5	LB111443BL		1	0.292	0.119	10/12/2020	12:03
6	L4214-07		1	0.724	0.559	10/12/2020	12:06
7	L4214-08		1	1.320	1.167	10/12/2020	12:08
8	CCV2	10	1	9.930	9.944	10/12/2020	12:10
9	CCB2		1	0.298	0.125	10/12/2020	12:12