



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111444

| Reagent/Standard                        | Lot/Log # |
|-----------------------------------------|-----------|
| TURBIDITY LOD STD, 0.5NTU               | WP87441   |
| Turbidity Calibration std, 0NTU         | WP87440   |
| Turbidity Calibration std, 20NTU        | WP87436   |
| Turbidity Calibration std, 40NTU        | WP87435   |
| Turbidity Calibration std, 1NTU         | WP87439   |
| Turbidity Calibration std, 5NTU         | WP87438   |
| Turbidity LOQ std, 1.0NTU               | WP87442   |
| Turbidity Calibration - CCV std, 10 NTU | WP87437   |

Intercept: 0.1753

Slope: 0.981

Regression: 0.9998

| Seq | Lab ID | True Value (NTU) | Dilution | Reading | Anal Date  | Anal Time |
|-----|--------|------------------|----------|---------|------------|-----------|
| 1   | CAL1   | 0                | 1        | 0.251   | 10/12/2020 | 11:40     |
| 2   | CAL2   | 1                | 1        | 1.220   | 10/12/2020 | 11:42     |
| 3   | CAL3   | 5                | 1        | 5.320   | 10/12/2020 | 11:44     |
| 4   | CAL4   | 10               | 1        | 9.920   | 10/12/2020 | 11:47     |
| 5   | CAL5   | 20               | 1        | 19.200  | 10/12/2020 | 11:49     |
| 6   | CAL6   | 40               | 1        | 39.700  | 10/12/2020 | 11:52     |



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111444

| Seq | Lab ID     | True Value<br>(NTU) | Dilution | Reading | Result<br>(NTU) | AnalDate   | Anal Time |
|-----|------------|---------------------|----------|---------|-----------------|------------|-----------|
| 1   | ICV        | 10                  | 1        | 10.200  | 10.219          | 10/12/2020 | 11:54     |
| 2   | ICB        |                     | 1        | 0.282   | 0.109           | 10/12/2020 | 11:56     |
| 3   | CCV1       | 10                  | 1        | 9.790   | 9.801           | 10/12/2020 | 11:58     |
| 4   | CCB1       |                     | 1        | 0.289   | 0.116           | 10/12/2020 | 12:01     |
| 5   | LB111444BL |                     | 1        | 0.292   | 0.119           | 10/12/2020 | 12:03     |
| 6   | L4214-10   |                     | 1        | 0.724   | 0.559           | 10/12/2020 | 12:06     |
| 7   | L4214-11   |                     | 1        | 1.320   | 1.167           | 10/12/2020 | 12:08     |
| 8   | CCV2       | 10                  | 1        | 9.930   | 9.944           | 10/12/2020 | 12:10     |
| 9   | CCB2       |                     | 1        | 0.298   | 0.125           | 10/12/2020 | 12:12     |