



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111446

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP87453
Residual chlorine ICV-LCS, 0.4PPM	WP87452
Chlorine Calibration std, 0.1ppm	WP87447
Chlorine Calibration std, 0.2ppm	WP87448
Chlorine Calibration std, 0.8ppm	WP87450
Chlorine Calibration std, 0.0ppm	WP87446
Chlorine Calibration std, 1.6ppm	WP87451
Residual Chlorine LOQ 0.1ppm	WP87454
Residual Chlorine Calibration and CCV std, 0	WP87449
Total Chlorine Powder Pillows	W2249

Intercept: 0.0086

Slope: 1.0705

Regression: 0.999913

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	10/12/2020	15:50
2	CAL2	0.1	1	0.000	0.110	0.110	10/12/2020	15:52
3	CAL3	0.2	1	0.000	0.230	0.230	10/12/2020	15:54
4	CAL4	0.4	1	0.000	0.450	0.450	10/12/2020	15:56
5	CAL5	0.8	1	0.000	0.860	0.860	10/12/2020	15:58
6	CAL6	1.6	1	0.000	1.720	1.720	10/12/2020	16:00



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111446

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4400	0.4400	0.4030	10/12/2020	16:03
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	10/12/2020	16:05
3	CCV1			0.4	1	0.0000	0.4300	0.4300	0.3940	10/12/2020	16:07
4	CCB1				1	0.0000	0.0000	0.0000	-0.0080	10/12/2020	16:10
5	LB111446BL	50	50		1	0.0000	0.0000	0.0000	-0.0080	10/12/2020	16:12
6	LB111446BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.3840	10/12/2020	16:14
7	L4214-07	50	50		1	0.0000	0.0700	0.0700	0.0570	10/12/2020	16:17
8	L4214-08	50	50		1	0.0000	0.1200	0.1200	0.1040	10/12/2020	16:19
9	CCV2			0.4	1	0.0000	0.4400	0.4400	0.4030	10/12/2020	16:22
10	CCB2				1	0.0000	0.0000	0.0000	-0.0080	10/12/2020	16:24