



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111453

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/12/2020	17:05
2	CAL2	1	Water	NA	NA	20.1	7.00	10/12/2020	17:06
3	CAL3	1	Water	NA	NA	20.1	10.03	10/12/2020	17:08
4	ICV	1	Water	NA	NA	20.1	6.99	10/12/2020	17:10
5	CCV1	1	Water	NA	NA	20.1	2.01	10/12/2020	17:11
6	L4217-06	1	Solid	20.03	20	20.0	7.82	10/12/2020	17:12
7	L4372-02	1	Solid	20.04	20	20.2	7.49	10/12/2020	17:15
8	L4372-02DUP	1	Solid	20.03	20	20.3	7.50	10/12/2020	17:16
9	CCV2	1	Water	NA	NA	20.1	12.02	10/12/2020	17:19

- JB 111453

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4217 WorkList ID : 143214 Department : Wet-Chemistry Date : 10-12-2020 15:41:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2020	Solid	L4217-06	Corrosivity	Cool 4 deg C	PSEG05	O61	HR-01-101220	10/12/2020	9045D
10/12/2020	Solid	L4372-02	Corrosivity	Cool 4 deg C	PSEG01	O61	WC-HADLEY	10/12/2020	9045D

Date/Time 10-12-2020 15:50
Received by: JB wet-chem
Relinquished by: KS (wet-chem)

Date/Time 10-12-2020 17:30
Received by: KS (wet-chem)
Relinquished by: JB wet-chem