


LB111464

7471B

INSTRUMENT ID: CY1

 μ Abs.:

44103



Linear

A= 0.0000e+000

B= 2.2606e-004

C= -4.1774e-002

Rho= 0.9996250

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	91	0.000	91				
0.2	0.200	0.183	-0.017	996	0.0 %	996				
2.5	2.500	2.423	-0.077	10901	0.0 %	10901				
5.0	5.000	5.220	0.220	23275	0.0 %	23275				
7.5	7.500	7.467	-0.033	33217	0.0 %	33217				
10.0	10.000	9.928	-0.072	44103	0.0 %	44103				

LB111464 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Oct 2020 11:08:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.0000	7471B	PPB	13 Oct 2020 11:09:57	S	Std
0.2 - 1	50.2	996	-	0.2000	7471B	PPB	13 Oct 2020 11:12:15	S	Std
2.5 - 1	52.5	10901	-	2.5000	7471B	PPB	13 Oct 2020 11:14:32	S	Std
5.0 - 1	55	23275	-	5.0000	7471B	PPB	13 Oct 2020 11:16:49	S	Std
7.5 - 1	57.5	33217	-	7.5000	7471B	PPB	13 Oct 2020 11:19:05	S	Std
10.0 - 1	510	44103	-	10.0000	7471B	PPB	13 Oct 2020 11:23:54	S	Std
ICV54 - 1	ICV54	18360	4.1087	-	7471B	PPB	13 Oct 2020 11:26:31	U	SMPL
ICB54 - 1	ICB54	-54	-0.0540	-	7471B	PPB	13 Oct 2020 11:28:46	U	SMPL
CCV16 - 1	CCV16	23487	5.2677	-	7471B	PPB	13 Oct 2020 11:31:01	U	SMPL
CCB16 - 1	CCB16	-54	-0.0540	-	7471B	PPB	13 Oct 2020 11:33:16	U	SMPL
CRA - 1	CRA	976	0.1789	-	7471B	PPB	13 Oct 2020 11:35:32	U	SMPL
HighStd - 1	HighStd	44041	9.9142	-	7471B	PPB	13 Oct 2020 11:37:47	U	SMPL
ChkStd - 1	ChkStd	31332	7.0412	-	7471B	PPB	13 Oct 2020 11:40:03	U	SMPL
PB132304BL - 1	PBS	-94	-0.0630	-	7471B	PPB	13 Oct 2020 11:42:21	U	SMPL
PB132304BS - 1	LCSS	18339	4.1040	-	7471B	PPB	13 Oct 2020 11:44:38	U	SMPL
L4217-01 - 1	HR-01-101220	4015	0.8659	-	7471B	PPB	13 Oct 2020 11:46:54	U	SMPL
L4217-01DUP - 1	HR-01-101220DUP	3538	0.7580	-	7471B	PPB	13 Oct 2020 11:49:10	U	SMPL
L4217-01MS - 1	HR-01-101220MS	23791	5.3364	-	7471B	PPB	13 Oct 2020 11:51:28	U	SMPL
L4217-01MSD - 1	HR-01-101220MSD	20466	4.5848	-	7471B	PPB	13 Oct 2020 11:53:45	U	SMPL
L4367-01 - 1	PM-WCS-TS-001	113143	25.5355	-	7471B	PPB	13 Oct 2020 11:56:02	U	SMPL
CCV17 - 1	CCV17	22532	5.0518	-	7471B	PPB	13 Oct 2020 11:58:18	U	SMPL
CCB17 - 1	CCB17	-99	-0.0642	-	7471B	PPB	13 Oct 2020 12:00:36	U	SMPL
L4369-02 - 1	HD-1	3432	0.7341	-	7471B	PPB	13 Oct 2020 12:02:53	U	SMPL
L4369-04 - 1	HD-2	3851	0.8288	-	7471B	PPB	13 Oct 2020 12:05:09	U	SMPL
L4369-06 - 1	HD-3	3235	0.6895	-	7471B	PPB	13 Oct 2020 12:07:24	U	SMPL
L4369-08 - 1	HD-4	2221	0.4603	-	7471B	PPB	13 Oct 2020 12:09:39	U	SMPL
L4369-10 - 1	HD-5	3852	0.8290	-	7471B	PPB	13 Oct 2020 12:11:55	U	SMPL
L4370-02 - 1	343	13974	3.1172	-	7471B	PPB	13 Oct 2020 12:14:11	U	SMPL
L4372-01 - 1	WC-HADLEY	2056	0.4230	-	7471B	PPB	13 Oct 2020 12:16:28	U	SMPL
L4217-01LX5 - 1		804	0.1400	-	7471B	PPB	13 Oct 2020 12:28:10	U	SMPL
L4217-01A - 1		22309	5.0014	-	7471B	PPB	13 Oct 2020 12:30:27	U	SMPL
L4367-01DLX5 - 1		21916	4.9126	-	7471B	PPB	13 Oct 2020 12:32:42	U	SMPL
CCV18 - 1	CCV18	22003	4.9322	-	7471B	PPB	13 Oct 2020 12:34:58	U	SMPL
CCB18 - 1	CCB18	-48	-0.0526	-	7471B	PPB	13 Oct 2020 12:37:16	U	SMPL