



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111467

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP87494
Turbidity Calibration std, 0NTU	WP87493
Turbidity Calibration std, 20NTU	WP87489
Turbidity Calibration std, 80NTU	WP87487
Turbidity Calibration std, 40NTU	WP87488
Turbidity Calibration std, 200NTU	WP87486
Turbidity Calibration std, 1NTU	WP87492
Turbidity Calibration std, 5NTU	WP87491
Turbidity Calibration - CCV std, 10 NTU	WP87490
10 NTU Standard 500 ml	W2489

Intercept: 0.3174

Slope: 0.9674

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.265	10/13/2020	12:15
2	CAL2	1	1	1.200	10/13/2020	12:17
3	CAL3	5	1	5.280	10/13/2020	12:19
4	CAL4	10	1	9.790	10/13/2020	12:21
5	CAL5	20	1	19.800	10/13/2020	12:23
6	CAL6	40	1	39.200	10/13/2020	12:25
7	CAL7	80	1	77.600	10/13/2020	12:27



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB111467

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.319	10/13/2020	12:30
2	ICB		1	0.278	-0.041	10/13/2020	12:32
3	CCV1	10	1	9.490	9.482	10/13/2020	12:34
4	CCB1		1	0.292	-0.026	10/13/2020	12:36
5	LB111467BL		1	0.295	-0.023	10/13/2020	12:39
6	L4214-12		1	0.712	0.408	10/13/2020	12:42
7	L4384-01		1	62.700	64.485	10/13/2020	12:45
8	L4384-01DUP		1	64.800	66.656	10/13/2020	12:47
9	CCV2	10	1	9.920	9.926	10/13/2020	12:49
10	CCB2		1	0.299	-0.019	10/13/2020	12:51

LB111467
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-101320 WorkList ID : 143249 Department : Wet-Chemistry Date : 10-13-2020 11:47:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/14/2020	Water	L4384-01	Turbidity	Cool 4 deg C	LIRO01	O61	UDS-4-26	10/12/2020	SM2130 B

Date/Time 10/13/2020 11:50
 Received by: Am (wet-chem)
 Relinquished by: CP (SMI)

Date/Time 10/13/2020 13:00
 Received by: CP (SMI)
 Relinquished by: Am (wet-chem)