



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111468

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP87494
Turbidity Calibration std, 0NTU	WP87493
Turbidity Calibration std, 20NTU	WP87489
Turbidity Calibration std, 40NTU	WP87488
Turbidity Calibration std, 1NTU	WP87492
Turbidity Calibration std, 5NTU	WP87491
Turbidity Calibration - CCV std, 10 NTU	WP87490
10 NTU Standard 500 ml	W2489

Intercept: 0.2558

Slope: 0.9737

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.265	10/13/2020	12:15
2	CAL2	1	1	1.200	10/13/2020	12:17
3	CAL3	5	1	5.280	10/13/2020	12:19
4	CAL4	10	1	9.790	10/13/2020	12:21
5	CAL5	20	1	19.800	10/13/2020	12:23
6	CAL6	40	1	39.200	10/13/2020	12:25



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111468

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.315	10/13/2020	12:30
2	ICB		1	0.278	0.023	10/13/2020	12:32
3	CCV1	10	1	9.490	9.484	10/13/2020	12:34
4	CCB1		1	0.292	0.037	10/13/2020	12:36
5	LB111468BL		1	0.295	0.040	10/13/2020	12:39
6	L4214-09		1	0.712	0.469	10/13/2020	12:42
7	CCV2	10	1	9.920	9.925	10/13/2020	12:49
8	CCB2		1	0.299	0.044	10/13/2020	12:51