



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111480

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP87544
Residual chlorine ICV-LCS, 0.4PPM	WP87543
Chlorine Calibration std, 0.1ppm	WP87538
Chlorine Calibration std, 0.2ppm	WP87539
Chlorine Calibration std, 0.8ppm	WP87541
Chlorine Calibration std, 0.0ppm	WP87537
Chlorine Calibration std, 1.6ppm	WP87542
Residual Chlorine Calibration and CCV std, 0	WP87540
Total Chlorine Powder Pillows	W2249

Intercept: 0.0066

Slope: 1.0582

Regression: 0.999951

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	10/13/2020	15:40
2	CAL2	0.1	1	0.000	0.110	0.110	10/13/2020	15:42
3	CAL3	0.2	1	0.000	0.230	0.230	10/13/2020	15:44
4	CAL4	0.4	1	0.000	0.430	0.430	10/13/2020	15:47
5	CAL5	0.8	1	0.000	0.850	0.850	10/13/2020	15:49
6	CAL6	1.6	1	0.000	1.700	1.700	10/13/2020	15:52



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB111480

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4400	0.4400	0.4100	10/13/2020	15:55
2	ICB				1	0.0000	0.0000	0.0000	-0.0060	10/13/2020	15:57
3	CCV1			0.4	1	0.0000	0.4500	0.4500	0.4190	10/13/2020	15:59
4	CCB1				1	0.0000	0.0000	0.0000	-0.0060	10/13/2020	16:02
5	LB111480BL	50	50		1	0.0000	0.0000	0.0000	-0.0060	10/13/2020	16:04
6	LB111480BS	50	50	0.4	1	0.0000	0.4400	0.4400	0.4100	10/13/2020	16:06
7	L4214-09	50	50		1	0.0000	0.0600	0.0600	0.0500	10/13/2020	16:09
8	CCV2			0.4	1	0.0000	0.4300	0.4300	0.4000	10/13/2020	16:11
9	CCB2				1	0.0000	0.0000	0.0000	-0.0060	10/13/2020	16:14