



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/14/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111483

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L4226-07	TR-08-101320	1	1.14	9.93	8.63	85.2	
L4226-08	TR-08-101320	2	1.14	9.8	8.67	87.0	
L4226-09	TR-08-101320-1	3	1.13	9.76	8.53	85.7	
L4226-13	TR-04-101320	4	1.16	9.8	8.84	88.9	
L4226-14	TR-04-101320	5	1.13	9.7	8.5	86.0	
L4226-15	TR-04-101320-1	6	1.11	9.88	8.77	87.3	
L4380-01	A-CONCRETE	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4380-03	A-CONTAINMENT	8	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4380-05	B-BRICK-1	9	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4380-07	B-BRICK-2	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4380-09	B-CONCRETE	11	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4380-11	C-CONCRETE	12	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4383-01	BB-SOIL	13	1.14	9.63	8.29	84.2	
L4383-02	BB-SOIL	14	1.13	9.72	8.18	82.1	
L4388-01	BG-1	15	1.14	9.79	7.15	69.5	
L4388-02	BG-2	16	1.15	9.95	8.24	80.6	
L4388-03	BG-3	17	1.12	9.75	8.53	85.9	
L4389-01	BG-1	18	1.17	9.98	8.33	81.3	
L4389-02	BG-2	19	1.13	9.61	8.09	82.1	
L4389-03	BG-3	20	1.15	9.92	7.73	75.0	
L4390-01	BG-1	21	1.11	9.68	8.35	84.5	
L4390-02	BG-2	22	1.13	9.65	8.09	81.7	
L4390-03	BG-3	23	1.1	9.89	8.29	81.8	
L4391-01	S1	24	1.18	9.55	7.89	80.2	
L4391-02	S2	25	1.13	9.81	8.09	80.2	
L4391-03	S3	26	1.18	9.96	7.85	76.0	
L4391-04	D101320	27	1.15	9.97	9.00	89.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$