


LB111490

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5386e-004

C= -8.4122e-002

Rho= 0.9999178

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.034	-0.034	197	0.000	197				
0.05	0.050					400				
0.20	0.200	0.188	-0.012	1073	0.0 %	1073				
2.50	2.500	2.499	-0.001	10174	0.0 %	10174				
5.00	5.000	5.100	0.100	20423	0.0 %	20423				
7.50	7.500	7.491	-0.009	29841	0.0 %	29841				
10.0	10.000	9.956	-0.044	39549	0.0 %	39549				

0100
116
00200

LB111490
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	197	-		0 ISM02.4	PPB	10/14/2020 9:08	Std
	0.2 S0.2	1073	-		0.2 ISM02.4	PPB	10/14/2020 9:10	Std
	2.5 S2.5	10174	-		2.5 ISM02.4	PPB	10/14/2020 9:12	Std
	5 S5.0	20423	-		5 ISM02.4	PPB	10/14/2020 9:15	Std
	7.5 S7.5	29841	-		7.5 ISM02.4	PPB	10/14/2020 9:17	Std
	10 S10.0	39549	-		10 ISM02.4	PPB	10/14/2020 9:19	Std
ICV	ICV	16231	4.0362	-	ISM02.4	PPB	10/14/2020 9:22	SMPL
ICB	ICB	168	-0.0415	-	ISM02.4	PPB	10/14/2020 9:24	SMPL
CCV019	CCV019	20597	5.1446	-	ISM02.4	PPB	10/14/2020 9:28	SMPL
CCB019	CCB019	156	-0.0445	-	ISM02.4	PPB	10/14/2020 9:30	SMPL
PB132345BL	PBS345	57	-0.0697	-	ISM02.4	PPB	10/14/2020 9:33	SMPL
L4268-01	MBEQC3	21667	5.4162	-	ISM02.4	PPB	10/14/2020 9:35	SMPL
L4268-02	MBEQC3D	20028	5.0001	-	ISM02.4	PPB	10/14/2020 9:37	SMPL
L4268-03	MBEQC3S	30566	7.6753	-	ISM02.4	PPB	10/14/2020 9:39	SMPL
L4268-04	MBEQ55	4618	1.0882	-	ISM02.4	PPB	10/14/2020 9:42	SMPL
L4268-05	MBEQ56	11989	2.9594	-	ISM02.4	PPB	10/14/2020 9:44	SMPL
L4268-06	MBEQ57	7725	1.8769	-	ISM02.4	PPB	10/14/2020 9:46	SMPL
L4268-07	MBEQ58	10577	2.6009	-	ISM02.4	PPB	10/14/2020 9:48	SMPL
L4268-08	MBEQB9	12643	3.1254	-	ISM02.4	PPB	10/14/2020 9:51	SMPL
L4268-09	MBEQC0	21418	5.353	-	ISM02.4	PPB	10/14/2020 9:53	SMPL
L4268-10	MBEQC1	32863	8.2584	-	ISM02.4	PPB	10/14/2020 9:55	SMPL
L4268-11	MBEQC5	25929	6.4981	-	ISM02.4	PPB	10/14/2020 9:58	SMPL
L4268-12	MBEQC6	9821	2.409	-	ISM02.4	PPB	10/14/2020 10:00	SMPL
L4268-13	MBEQC8	1988	0.4205	-	ISM02.4	PPB	10/14/2020 10:02	SMPL
L4268-14	MBEQC9	1671	0.3401	-	ISM02.4	PPB	10/14/2020 10:04	SMPL
L4268-15	MBEQ59	4439	1.0427	-	ISM02.4	PPB	10/14/2020 10:07	SMPL
L4268-16	MBEQ60	16428	4.0862	-	ISM02.4	PPB	10/14/2020 10:09	SMPL
L4268-17	MBEQ61	5623	1.3433	-	ISM02.4	PPB	10/14/2020 10:11	SMPL
L4268-18	MBEQ62	5135	1.2194	-	ISM02.4	PPB	10/14/2020 10:13	SMPL
L4268-19	MBEQ63	4433	1.0412	-	ISM02.4	PPB	10/14/2020 10:16	SMPL
CCV020	CCV020	20683	5.1664	-	ISM02.4	PPB	10/14/2020 10:18	SMPL
CCB020	CCB020	144	-0.0476	-	ISM02.4	PPB	10/14/2020 10:20	SMPL
L4268-20	MBEQ64	6767	1.6337	-	ISM02.4	PPB	10/14/2020 10:23	SMPL
L4268-21	MBEQ65	1154	0.2088	-	ISM02.4	PPB	10/14/2020 10:25	SMPL
L4268-22	MBEQ66	6786	1.6385	-	ISM02.4	PPB	10/14/2020 10:27	SMPL
PB132346BL	PBS346	0	-0.0841	-	ISM02.4	PPB	10/14/2020 10:29	SMPL
L4320-01	MBEQE6	15666	3.8928	-	ISM02.4	PPB	10/14/2020 10:32	SMPL
L4320-02	MBEQE7	6680	1.6116	-	ISM02.4	PPB	10/14/2020 10:34	SMPL
L4320-03	MBEQE8	1413	0.2746	-	ISM02.4	PPB	10/14/2020 10:36	SMPL
L4320-04	MBEQE9	6047	1.4509	-	ISM02.4	PPB	10/14/2020 10:38	SMPL
L4320-05	MBEQE9D	6333	1.5236	-	ISM02.4	PPB	10/14/2020 10:43	SMPL
L4320-06	MBEQE9S	15316	3.8039	-	ISM02.4	PPB	10/14/2020 10:46	SMPL
L4320-07	MBEQF0	15511	3.8534	-	ISM02.4	PPB	10/14/2020 10:48	SMPL
L4320-08	MBEQF1	31787	7.9852	-	ISM02.4	PPB	10/14/2020 10:50	SMPL
L4320-09	MBEQG3	28601	7.1764	-	ISM02.4	PPB	10/14/2020 10:53	SMPL
L4320-11	MBEQD7	7100	1.7183	-	ISM02.4	PPB	10/14/2020 10:55	SMPL

LB111490
INSTRUMENT ID : CV1

L4320-12	MBEQD8	8828	2.1569 -	ISM02.4	PPB	10/14/2020 10:57 SMPL
L4320-13	MBEQD9	19780	4.9372 -	ISM02.4	PPB	10/14/2020 10:59 SMPL
L4320-14	MBEQE0	5592	1.3354 -	ISM02.4	PPB	10/14/2020 11:02 SMPL
L4320-15	MBEQE1	18359	4.5764 -	ISM02.4	PPB	10/14/2020 11:04 SMPL
L4320-16	MBEQE2	24314	6.0881 -	ISM02.4	PPB	10/14/2020 11:06 SMPL
L4320-17	MBEQE3	1571	0.3147 -	ISM02.4	PPB	10/14/2020 11:08 SMPL
CCV021	CCV021	18335	4.5703 -	ISM02.4	PPB	10/14/2020 11:11 SMPL
CCB021	CCB021	116	-0.0547 -	ISM02.4	PPB	10/14/2020 11:13 SMPL
L4320-18	MBEQE4	1330	0.2535 -	ISM02.4	PPB	10/14/2020 11:15 SMPL
L4320-19	MBEQE5	1468	0.2885 -	ISM02.4	PPB	10/14/2020 11:17 SMPL
PB132347BL	PBS347	296	-0.009 -	ISM02.4	PPB	10/14/2020 11:21 SMPL
L4323-01	MJLX60	773	0.1121 -	ISM02.4	PPB	10/14/2020 11:23 SMPL
L4323-02	MJLX60D	858	0.1337 -	ISM02.4	PPB	10/14/2020 11:29 SMPL
L4323-03	MJLX60S	12450	3.0764 -	ISM02.4	PPB	10/14/2020 11:31 SMPL
L4323-04	MJLX61	508	0.0448 -	ISM02.4	PPB	10/14/2020 11:33 SMPL
CCV022	CCV022	21407	5.3502 -	ISM02.4	PPB	10/14/2020 11:38 SMPL
CCB022	CCB022	215	-0.0295 -	ISM02.4	PPB	10/14/2020 11:40 SMPL