

L3111494


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

36149



A= 0.0000e+000

B= 2.7343e-004

C= -1.9007e-002

Rho= 0.9994176

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	104	0.000	104				
0.2	0.200	0.195	-0.005	781	0.0 %	781				
2.5	2.500	2.508	0.008	9241	0.0 %	9241				
5.0	5.000	4.875	-0.125	17900	0.0 %	17900				
7.5	7.500	7.748	0.248	28404	0.0 %	28404				
10.0	10.000	9.865	-0.135	36149	0.0 %	36149				

LB111494 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Oct 2020 12:31:34

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7470A	PPB	14 Oct 2020 12:31:53	S	Std
0.2 - 1	50.2	781	-	0.2000	7470A	PPB	14 Oct 2020 12:34:10	S	Std
2.5 - 1	52.5	9241	-	2.5000	7470A	PPB	14 Oct 2020 12:36:27	S	Std
5.0 - 1	55	17900	-	5.0000	7470A	PPB	14 Oct 2020 12:38:44	S	Std
7.5 - 1	57.5	28404	-	7.5000	7470A	PPB	14 Oct 2020 12:41:00	S	Std
10.0 - 1	59	36149	-	10.0000	7470A	PPB	14 Oct 2020 12:45:49	S	Std
ICV55 - 1	ICV55	13679	3.7213	-	7470A	PPB	14 Oct 2020 12:48:31	U	SMPL
ICB55 - 1	ICB55	-74	-0.0392	-	7470A	PPB	14 Oct 2020 12:50:46	U	SMPL
CCV23 - 1	CCV23	17863	4.8653	-	7470A	PPB	14 Oct 2020 12:53:01	U	SMPL
CCB23 - 1	CCB23	-70	-0.0381	-	7470A	PPB	14 Oct 2020 12:55:17	U	SMPL
CRA - 1	CRA	736	0.1822	-	7470A	PPB	14 Oct 2020 12:57:32	U	SMPL
HighStd - 1	HighStd	36828	10.0509	-	7470A	PPB	14 Oct 2020 12:59:47	U	SMPL
ChkStd - 1	ChkStd	26600	7.2543	-	7470A	PPB	14 Oct 2020 13:02:02	U	SMPL
PB132341BL - 1	PBW	53	-0.0045	-	7470A	PPB	14 Oct 2020 13:08:18	U	SMPL
PB132341BS - 1	LCSW	12529	3.4068	-	7470A	PPB	14 Oct 2020 13:10:34	U	SMPL
L4217-06 - 1	HR-01-101220	-23	-0.0253	-	7470A	PPB	14 Oct 2020 13:12:50	U	SMPL
L4217-06DUP - 1	HR-01-101220DUP	11	-0.0160	-	7470A	PPB	14 Oct 2020 13:15:06	U	SMPL
L4217-06MS - 1	HR-01-101220MS	13810	3.7571	-	7470A	PPB	14 Oct 2020 13:17:23	U	SMPL
L4217-06MSD - 1	HR-01-101220MSD	13585	3.6956	-	7470A	PPB	14 Oct 2020 13:19:40	U	SMPL
L4367-02 - 1	PM-WCS-TS-001	-49	-0.0324	-	7470A	PPB	14 Oct 2020 13:21:55	U	SMPL
CCV24 - 1	CCV24	17393	4.7368	-	7470A	PPB	14 Oct 2020 13:24:10	U	SMPL
CCB24 - 1	CCB24	-70	-0.0381	-	7470A	PPB	14 Oct 2020 13:26:25	U	SMPL
L4368-01 - 1	10-12A	19	-0.0138	-	7470A	PPB	14 Oct 2020 13:28:41	U	SMPL
L4368-02 - 1	10-12B	0	-0.0190	-	7470A	PPB	14 Oct 2020 13:30:56	U	SMPL
L4368-03 - 1	10-12C	21	-0.0133	-	7470A	PPB	14 Oct 2020 13:33:11	U	SMPL
PB132354BL - 1	PBW	111	0.0113	-	7470A	PPB	14 Oct 2020 13:35:27	U	SMPL
PB132354BS - 1	LCSW	13651	3.7136	-	7470A	PPB	14 Oct 2020 13:37:43	U	SMPL
L4214-07 - 1	LOD-MDL-WATER-01-QT4-2020	258	0.0515	-	7470A	PPB	14 Oct 2020 13:42:34	U	SMPL
L4214-08 - 1	LOQ-WATER-02-QT4-2020	728	0.1801	-	7470A	PPB	14 Oct 2020 13:47:26	U	SMPL
L4381-07 - 1	COMP-FRAC-TANKS	1047	0.2673	-	7470A	PPB	14 Oct 2020 13:54:34	U	SMPL
L4381-07DUP - 1	COMP-FRAC-TANKSDUP	1030	0.2626	-	7470A	PPB	14 Oct 2020 13:56:51	U	SMPL
L4381-07MS - 1	COMP-FRAC-TANKSMS	13379	3.6392	-	7470A	PPB	14 Oct 2020 14:01:46	U	SMPL
CCV25 - 1	CCV25	16879	4.5962	-	7470A	PPB	14 Oct 2020 14:04:01	U	SMPL
CCB25 - 1	CCB25	-51	-0.0330	-	7470A	PPB	14 Oct 2020 14:06:17	U	SMPL
L4381-07MSD - 1	COMP-FRAC-TANKSMSD	14608	3.9753	-	7470A	PPB	14 Oct 2020 14:08:32	U	SMPL
PB132341TB - 1	PB132302TB	-66	-0.0371	-	7470A	PPB	14 Oct 2020 14:10:47	U	SMPL
L4217-06LX5 - 1		40	-0.0081	-	7470A	PPB	14 Oct 2020 14:13:04	U	SMPL
L4217-06A - 1		13343	3.6294	-	7470A	PPB	14 Oct 2020 14:15:19	U	SMPL
L4381-07LX5 - 1		187	0.0321	-	7470A	PPB	14 Oct 2020 14:17:34	U	SMPL
L4381-07A - 1		15364	4.1820	-	7470A	PPB	14 Oct 2020 14:19:51	U	SMPL
CCV26 - 1	CCV26	17442	4.7502	-	7470A	PPB	14 Oct 2020 14:22:08	U	SMPL
CCB26 - 1	CCB26	-8	-0.0212	-	7470A	PPB	14 Oct 2020 14:27:00	U	SMPL