



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111496

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP87561
Turbidity Calibration std, 20NTU	WP87557
Turbidity Calibration std, 40NTU	WP87556
Turbidity Calibration std, 1NTU	WP87560
Turbidity Calibration std, 5NTU	WP87559
Turbidity Calibration - CCV std, 10 NTU	WP87558
10 NTU Standard 500 ml	W2489

Intercept: 0.2388

Slope: 0.9746

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.260	10/14/2020	11:20
2	CAL2	1	1	1.180	10/14/2020	11:22
3	CAL3	5	1	5.140	10/14/2020	11:24
4	CAL4	10	1	9.920	10/14/2020	11:26
5	CAL5	20	1	19.800	10/14/2020	11:28
6	CAL6	40	1	39.200	10/14/2020	11:30



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111496

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.426	10/14/2020	11:33
2	ICB		1	0.280	0.042	10/14/2020	11:35
3	CCV1	10	1	9.720	9.728	10/14/2020	11:37
4	CCB1		1	0.285	0.047	10/14/2020	11:40
5	LB111496BL		1	0.290	0.053	10/14/2020	11:42
6	L4398-01		1	8.920	8.907	10/14/2020	11:45
7	L4398-01DUP		1	9.320	9.318	10/14/2020	11:48
8	L4399-01		1	9.170	9.164	10/14/2020	11:50
9	CCV2	10	1	9.880	9.892	10/14/2020	11:52
10	CCB2		1	0.293	0.056	10/14/2020	11:55

LB111496
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-101420 WorkList ID : 143300 Department : Wet-Chemistry Date : 10-14-2020 10:58:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Water	L4398-01	Turbidity	Cool 4 deg C	LIRO01	P11	UDS-4-40	10/13/2020	SM2130 B
10/15/2020	Water	L4399-01	Turbidity	Cool 4 deg C	LIRO01	P11	UDS-6-16A	10/13/2020	SM2130 B

Date/Time 10/14/2020 11:05
 Received by: AM (wet chem)
 Relinquished by: CP (sm)

Date/Time 10/14/2020 12:20
 Received by: CP (sm)
 Relinquished by: AM (wet chem)