



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111497

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	18.5	4.01	10/14/2020	12:35
2	CAL2	1	Water	NA	NA	18.5	7.01	10/14/2020	12:37
3	CAL3	1	Water	NA	NA	18.5	10.03	10/14/2020	12:39
4	ICV	1	Water	NA	NA	18.5	6.98	10/14/2020	12:42
5	CCV1	1	Water	NA	NA	18.6	2.01	10/14/2020	12:44
6	L4400-01	1	Water	NA	NA	18.2	5.11	10/14/2020	12:47
7	L4400-01DUP	1	Water	NA	NA	18.2	5.12	10/14/2020	12:49
8	L4400-02	1	Water	NA	NA	18.0	4.70	10/14/2020	12:52
9	CCV2	1	Water	NA	NA	18.6	12.02	10/14/2020	12:55

LB111497
WORKLIST(Hardcopy Internal Chain)

WorkList Name : pHwater-101420

WorkList ID : 143301

Department : Wet-Chemistry

Date : 10-14-2020 12:11:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/13/2020	Water	L4400-01	pH	Cool 4 deg C	STER16	P11	MAIN-CELL	10/13/2020	9040C
10/13/2020	Water	L4400-02	pH	Cool 4 deg C	STER16	P11	EX-CELL	10/13/2020	9040C

Date/Time 10/14/2020 12:20
Received by: AMI wet-chem
Relinquished by: CP (Sm)

Date/Time 10/14/2020 13:10
Received by: CP (Sm)
Relinquished by: AMI wet-chem