



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111500

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/14/2020	10:58
2	L4381-13		1	8	100.00	>212.0	>212.0	10/14/2020	11:25
3	L4381-13DUP		1	8	100.00	>212.0	>212.0	10/14/2020	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111500

WorkList Name : FP101420 WorkList ID : 143285 Department : Wet-Chemistry Date : 10-14-2020 09:26:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Water	L4381-13	Flash Point	Cool 4 deg C	PSEG01		DRUM-COMP	10/13/2020	1010A

Date/Time 10/14/20 09:30
 Received by: KS (Wetchem)
 Relinquished by: RJ (Ext. 196)

Date/Time 10/14/20 11:55
 Received by: RJ (Ext. 196)
 Relinquished by: KS (Wetchem)