



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111501

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/14/2020	14:38
2	CAL2	1	Water	NA	NA	20.2	7.01	10/14/2020	14:39
3	CAL3	1	Water	NA	NA	20.2	10.01	10/14/2020	14:40
4	ICV	1	Water	NA	NA	20.2	6.99	10/14/2020	14:42
5	CCV1	1	Water	NA	NA	20.1	2.01	10/14/2020	14:44
6	L4403-01	1	Water	NA	NA	19.4	11.72	10/14/2020	14:46
7	L4403-01DUP	1	Water	NA	NA	19.5	11.74	10/14/2020	14:48
8	L4403-05	1	Water	NA	NA	19.5	7.76	10/14/2020	14:50
9	CCV2	1	Water	NA	NA	20.1	12.02	10/14/2020	14:52

LB11501

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH-2-101420

WorkList ID : 143308

Department : Wet-Chemistry

Date : 10-14-2020 13:05:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/14/2020	Water	L4403-01 E	pH	Cool 4 deg C	TULL02	P11	SHELL-L15-L14-NORTH	10/14/2020	9040C
10/14/2020	Water	L4403-05 D	pH	Cool 4 deg C	TULL02	P11	MH-N	10/14/2020	9040C

Date/Time 10/14/2020 14:30
Received by: ArLwet-chem
Relinquished by: CL (Sm)

Date/Time 10/14/2020 14:55
Received by: CL (Sm)
Relinquished by: ArLwet-chem