



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111510

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/15/2020	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	10/15/2020	08:16
3	CAL3	1	Water	NA	NA	20.2	10.03	10/15/2020	08:17
4	ICV	1	Water	NA	NA	20.2	7.00	10/15/2020	08:19
5	CCV1	1	Water	NA	NA	20.3	2.01	10/15/2020	08:20
6	L4223-08	1	Solid	20.03	20	20.3	7.72	10/15/2020	08:21
7	L4227-04	1	Solid	20.04	20	19.9	9.23	10/15/2020	08:22
8	L4393-01	1	Solid	20.02	20	19.4	6.84	10/15/2020	08:23
9	L4393-02	1	Solid	20.03	20	19.5	6.63	10/15/2020	08:25
10	L4394-01	1	Solid	20.02	20	19.4	5.75	10/15/2020	08:26
11	L4395-01	1	Solid	20.03	20	19.2	6.10	10/15/2020	08:28
12	L4395-03	1	Solid	20.04	20	19.2	6.17	10/15/2020	08:29
13	L4404-02	1	Solid	20.02	20	21.1	9.13	10/15/2020	08:30
14	L4406-08	1	Solid	20.03	20	21.5	8.75	10/15/2020	08:33
15	L4406-13	1	Solid	20.04	20	21.6	7.70	10/15/2020	08:35
16	CCV2	1	Water	NA	NA	20.3	12.02	10/15/2020	08:36
17	L4406-13DUP	1	Solid	20.03	20	21.7	7.71	10/15/2020	08:37
18	CCV3	1	Water	NA	NA	20.1	2.01	10/15/2020	08:39

WB 111510

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4395 WorkList ID : 143315 Department : Wet-Chemistry Date : 10-15-2020 08:07:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/14/2020	Solid	L4223-08	Corrosivity	Cool 4 deg C	PSEG05	O12	PL-01-101420	10/14/2020	9045D
10/14/2020	Solid	L4227-04	Corrosivity	Cool 4 deg C	PSEG05	O12	BU-01-101420	10/14/2020	9045D
10/14/2020	Solid	L4393-01	Corrosivity	Cool 4 deg C	PSEG01	P11	OILY-DEBRIS-DRUM	10/14/2020	9045D
10/14/2020	Solid	L4393-02	Corrosivity	Cool 4 deg C	PSEG01	P11	OILY-STONE-DRUMS	10/14/2020	9045D
10/14/2020	Solid	L4394-01	Corrosivity	Cool 4 deg C	PSEG01	O12	BLUE-DRUM	10/14/2020	9045D
10/14/2020	Solid	L4395-01	Corrosivity	Cool 4 deg C	PSEG01	P11	OILY-DEBRIS-DRUMS-(1-4)	10/14/2020	9045D
10/14/2020	Solid	L4395-03	Corrosivity	Cool 4 deg C	PSEG01	P11	OILY-STONE-DRUMS-(1-2)	10/14/2020	9045D
10/14/2020	Solid	L4404-02	Corrosivity	Cool 4 deg C	TULL02	P11	SHELL-L15-L14-NORTH	10/14/2020	9045D
10/14/2020	Solid	L4406-08	Corrosivity	Cool 4 deg C	PSEG05	O12	JC-02-101420-A	10/14/2020	9045D
10/14/2020	Solid	L4406-13	Corrosivity	Cool 4 deg C	PSEG05	O12	JC-02-101420-B	10/14/2020	9045D

Date/Time 10-15-2020 8:10
 Received by: JB [Signature]
 Relinquished by: JB [Signature]

Date/Time 10-13-2020 11:00
 Received by: JB [Signature]
 Relinquished by: JB [Signature]