



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111514

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/15/2020	08:15
2	CAL2	1	Water	NA	NA	20.1	7.00	10/15/2020	08:16
3	CAL3	1	Water	NA	NA	20.1	10.03	10/15/2020	08:16
4	ICV	1	Water	NA	NA	20.1	7.00	10/15/2020	08:19
5	CCV1	1	Water	NA	NA	20.1	2.02	10/15/2020	08:20
6	L4392-17	1	Water	NA	NA	20.3	7.04	10/15/2020	08:21
7	L4392-17DUP	1	Water	NA	NA	20.2	7.05	10/15/2020	08:22
8	CCV2	1	Water	NA	NA	20.1	12.02	10/15/2020	08:24

VB 111514

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14404 WorkList ID : 143313 Department : Wet-Chemistry Date : 10-15-2020 08:05:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/14/2020	Water	L4392-17	pH	Cool 4 deg C	PSEG01	O12	HACK-MH	10/14/2020	9040C

Date/Time 10-15-2020 8:10
Received by: Jd (wet)
Relinquished by: Jd (wet) Chem

Date/Time 10-15-2020 9:10
Received by: Jd (wet)
Relinquished by: Jd (wet) Chem