


LB111517

245.1

INSTRUMENT ID: CV4



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	76	0.000	76				
0.05	0.050	0.058	0.008	245	0.0 %	245	0			
0.2	0.200	0.188	-0.012	674	0.0 %	674				
2.5	2.500	2.365	-0.135	7878	0.0 %	7878				
5.0	5.000	5.132	0.132	17034	0.0 %	17034				
7.5	7.500	7.661	0.161	25401	0.0 %	25401				
10.0	10.000	9.838	-0.162	32605	0.0 %	32605				

LB111517 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Oct 2020 11:58:59

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	76	-	0.0000	245.1	PPB	15 Oct 2020 12:02:18	S	Std
0.05 - 1	50.85	245	-	0.0500	245.1	PPB	15 Oct 2020 12:04:35	S	Std
0.2 - 1	50.2	674	-	0.2000	245.1	PPB	15 Oct 2020 12:06:52	S	Std
2.5 - 1	52.5	7878	-	2.5000	245.1	PPB	15 Oct 2020 12:09:09	S	Std
5.0 - 1	55	17034	-	5.0000	245.1	PPB	15 Oct 2020 12:11:25	S	Std
7.5 - 1	57.5	25401	-	7.5000	245.1	PPB	15 Oct 2020 12:19:01	S	Std
10.0 - 1	510	32605	-	10.0000	245.1	PPB	15 Oct 2020 12:21:16	S	Std
ICV57 - 1	ICV57	13214	3.9778	-245.1		PPB	15 Oct 2020 12:29:15	U	SMPL
ICB57 - 1	ICB57	-78	-0.0394	-245.1		PPB	15 Oct 2020 12:31:30	U	SMPL
CCV31 - 1	CCV31	16917	5.0969	-245.1		PPB	15 Oct 2020 12:33:45	U	SMPL
CCB31 - 1	CCB31	-82	-0.0406	-245.1		PPB	15 Oct 2020 12:36:00	U	SMPL
CRA - 1	CRA	196	0.0435	-245.1		PPB	15 Oct 2020 12:38:16	U	SMPL
HighStd - 1	HighStd	32991	9.9549	-245.1		PPB	15 Oct 2020 12:40:31	U	SMPL
ChkStd - 1	ChkStd	24603	7.4198	-245.1		PPB	15 Oct 2020 12:42:48	U	SMPL
PB132394BL - 1	PBW	-80	-0.0400	-245.1		PPB	15 Oct 2020 12:45:05	U	SMPL
PB132394BS - 1	LCSW	13527	4.0724	-245.1		PPB	15 Oct 2020 12:50:26	U	SMPL
L4214-10 - 1	LOD-MDL-WATER-04-QT4-2020	207	0.0468	-245.1		PPB	15 Oct 2020 12:55:18	U	SMPL
L4214-11 - 1	LOQ-WATER-05-QT4-2020	695	0.1943	-245.1		PPB	15 Oct 2020 12:57:34	U	SMPL
L4350-07 - 1	14B-(1-4)-COMP	498	0.1347	-245.1		PPB	15 Oct 2020 12:59:51	U	SMPL
L4350-07DUP - 1	14B-(1-4)-COMPDUP	502	0.1359	-245.1		PPB	15 Oct 2020 13:04:41	U	SMPL
L4350-07MS - 1	14B-(1-4)-COMPMS	14395	4.3347	-245.1		PPB	15 Oct 2020 13:09:38	U	SMPL
CCV32 - 1	CCV32	16472	4.9624	-245.1		PPB	15 Oct 2020 13:11:53	U	SMPL
CCB32 - 1	CCB32	-81	-0.0403	-245.1		PPB	15 Oct 2020 13:14:09	U	SMPL
L4350-07MSD - 1	14B-(1-4)-COMPMSD	15399	4.6382	-245.1		PPB	15 Oct 2020 13:16:25	U	SMPL
L4371-07 - 1	14B-(1-4)-COMP	436	0.1160	-245.1		PPB	15 Oct 2020 13:18:40	U	SMPL
L4386-07 - 1	14B-(1-4)-COMP	213	0.0486	-245.1		PPB	15 Oct 2020 13:20:55	U	SMPL
L4350-07LX5 - 1		88	0.0108	-245.1		PPB	15 Oct 2020 13:23:11	U	SMPL
L4350-07A - 1		13722	4.1313	-245.1		PPB	15 Oct 2020 13:25:27	U	SMPL
CCV33 - 1	CCV33	16707	5.0335	-245.1		PPB	15 Oct 2020 13:27:43	U	SMPL
CCB33 - 1	CCB33	-62	-0.0345	-245.1		PPB	15 Oct 2020 13:29:59	U	SMPL