



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111518

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/15/2020	11:50
2	CAL2	1	Water	NA	NA	20.1	7.00	10/15/2020	11:51
3	CAL3	1	Water	NA	NA	20.1	10.03	10/15/2020	11:52
4	ICV	1	Water	NA	NA	20.2	7.01	10/15/2020	11:53
5	CCV1	1	Water	NA	NA	20.2	2.02	10/15/2020	11:55
6	L4415-01	1	Water	NA	NA	14.3	6.84	10/15/2020	12:00
7	L4415-01DUP	1	Water	NA	NA	14.4	6.85	10/15/2020	12:01
8	L4418-02	1	Water	NA	NA	14.2	9.84	10/15/2020	12:05
9	CCV2	1	Water	NA	NA	20.1	12.02	10/15/2020	12:06

VB 111518

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4418 W WorkList ID : 143348 Department : Wet-Chemistry Date : 10-15-2020 11:39:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/14/2020	Water	L4415-01	Corrosivity	Cool 4 deg C	TETR06	P12	CALTANK1-20201014	10/14/2020	9040C
10/14/2020	Water	L4418-02	Corrosivity	Cool 4 deg C	ARCA05	P12	PB-2020-WWW-000-02	10/14/2020	9040C

Date/Time 10:15-2020 11:40
 Received by: Jp Wet-chem
 Relinquished by: Jp

Date/Time 10:15-2020 12:10
 Received by: Jp
 Relinquished by: Jp Wet-chem