



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111523

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP87586
Turbidity Calibration std, 20NTU	WP87582
Turbidity Calibration std, 80NTU	WP87580
Turbidity Calibration std, 40NTU	WP87581
Turbidity Calibration std, 200NTU	WP87579
Turbidity Calibration std, 1NTU	WP87585
Turbidity Calibration std, 5NTU	WP87584
Turbidity Calibration - CCV std, 10 NTU	WP87583
10 NTU Standard 500 ml	W2489

Intercept: 0.1305

Slope: 0.9667

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.255	10/15/2020	12:35
2	CAL2	1	1	1.190	10/15/2020	12:37
3	CAL3	5	1	5.330	10/15/2020	12:39
4	CAL4	10	1	9.640	10/15/2020	12:41
5	CAL5	20	1	18.900	10/15/2020	12:43
6	CAL6	40	1	38.800	10/15/2020	12:45
7	CAL7	80	1	77.600	10/15/2020	12:48
8	CAL8	200	1	193.000	10/15/2020	12:50



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB111523

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.820	10.023	10/15/2020	12:53
2	ICB		1	0.282	0.157	10/15/2020	12:55
3	CCV1	10	1	9.640	9.837	10/15/2020	12:58
4	CCB1		1	0.292	0.167	10/15/2020	13:01
5	LB111523BL		1	0.278	0.153	10/15/2020	13:03
6	L4414-01		1	97.100	100.310	10/15/2020	13:05
7	L4414-01DUP		1	96.600	99.793	10/15/2020	13:08
8	CCV2	10	1	9.770	9.972	10/15/2020	13:10
9	CCB2		1	0.280	0.155	10/15/2020	13:12

LBUS23

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-101520 WorkList ID : 143339 Department : Wet-Chemistry Date : 10-15-2020 10:35:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2020	Water	L4414-01	Turbidity	Cool 4 deg C	LIRO01	P11	UDS-2-01	10/14/2020	SM2130 B

Date/Time 10/15/2020 12:10
Received by: ARM (wet-chem)
Relinquished by: AP (wet-chem)

Date/Time 10/15/2020 13:20
Received by: AP (wet-chem)
Relinquished by: ARM (wet-chem)