



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111537

Slope : 97.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	10/15/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.00	10/15/2020	17:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/15/2020	17:22
4	ICV	1	Water	NA	NA	20.1	7.00	10/15/2020	17:24
5	CCV1	1	Water	NA	NA	20.2	2.02	10/15/2020	17:25
6	L4411-04	1	Water	NA	NA	22.1	6.69	10/15/2020	17:26
7	L4412-01	1	Water	NA	NA	20.6	7.56	10/15/2020	17:28
8	L4412-02	1	Water	NA	NA	20.3	7.38	10/15/2020	17:30
9	L4412-03	1	Water	NA	NA	20.5	7.51	10/15/2020	17:32
10	L4412-04	1	Water	NA	NA	20.4	6.64	10/15/2020	17:33
11	L4412-05	1	Water	NA	NA	20.2	6.61	10/15/2020	17:34
12	L4412-05DUP	1	Water	NA	NA	20.3	6.61	10/15/2020	17:35
13	CCV2	1	Water	NA	NA	20.1	12.02	10/15/2020	17:36

111537

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4411 WorkList ID : 143364 Department : Wet-Chemistry Date : 10-15-2020 16:37:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Water	L4411-04	pH	Cool 4 deg C	PSEG01	O12	32697	10/15/2020	9040C
10/15/2020	Water	L4412-01	pH	Cool 4 deg C	PSEG01	O12	32699	10/15/2020	9040C
10/15/2020	Water	L4412-02	pH	Cool 4 deg C	PSEG01	O12	32698	10/15/2020	9040C
10/15/2020	Water	L4412-03	pH	Cool 4 deg C	PSEG01	O12	32700	10/15/2020	9040C
10/15/2020	Water	L4412-04	pH	Cool 4 deg C	PSEG01	O12	32701	10/15/2020	9040C
10/15/2020	Water	L4412-05	pH	Cool 4 deg C	PSEG01	O12	32702	10/15/2020	9040C

Date/Time 10-15-2020 16:40
 Received by: J. C. (Saw)
 Relinquished by: J. C. (Saw)

Date/Time 10-15-2020 18:30
 Received by: J. C. (Saw)
 Relinquished by: J. C. (Saw)