



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111538

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	10/15/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.00	10/15/2020	17:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/15/2020	17:22
4	ICV	1	Water	NA	NA	20.1	7.00	10/15/2020	17:24
5	CCV1	1	Water	NA	NA	20.2	2.02	10/15/2020	17:25
6	L4411-01	1	Solid	20.03	20	21.1	5.56	10/15/2020	17:38
7	L4411-02	1	Solid	20.03	20	21.5	7.20	10/15/2020	17:40
8	L4413-01	1	Solid	20.02	20	21.3	8.23	10/15/2020	17:42
9	L4413-01DUP	1	Solid	20.03	20	21.4	8.24	10/15/2020	17:43
10	CCV2	1	Water	NA	NA	20.3	12.02	10/15/2020	17:44

11538

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4411 S

WorkList ID : 143365

Department : Wet-Chemistry

Date : 10-15-2020 16:38:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Solid	L4411-01	pH	Cool 4 deg C	PSEG01	O12	18618	10/15/2020	9045D
10/15/2020	Solid	L4411-02	pH	Cool 4 deg C	PSEG01	O12	32696	10/15/2020	9045D
10/15/2020	Solid	L4413-01	pH	Cool 4 deg C	PSEG01	O12	30688	10/15/2020	9045D

Date/Time 10/15/2020 16:40
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 10/15/2020 18:30
 Received by: [Signature]
 Relinquished by: [Signature]