



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111539

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/15/2020	15:10
2	L4411-01		1	8	70.00	158.0	158.0	10/15/2020	18:15
3	L4411-01DUP		1	8	71.00	159.8	159.8	10/15/2020	18:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB111539

WorkList Name : FP101520-1 WorkList ID : 143363 Department : Wet-Chemistry Date : 10-15-2020 16:16:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2020	Solid	L4411-01	Flash Point	Cool 4 deg C	PSEG01	O12	18618	10/15/2020	1010A

Date/Time 10/15/2020 16:20
 Received by: KS (Wetchem)
 Relinquished by: RC (EXT LAB)

Date/Time 10/15/2020 18:45
 Received by: RC (EXT LAB)
 Relinquished by: KS (Wetchem)