



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111540

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/15/2020	15:10
2	L4392-18		1	8	100.00	>212.0	>212.0	10/15/2020	15:40
3	L4392-18DUP		1	8	100.00	>212.0	>212.0	10/15/2020	16:10
4	L4403-01		1	9	100.00	>212.0	>212.0	10/15/2020	16:35
5	L4403-05		1	8	100.00	>212.0	>212.0	10/15/2020	16:58
6	L4411-04		1	9	100.00	>212.0	>212.0	10/15/2020	17:20
7	L4418-02		1	8	100.00	>212.0	>212.0	10/15/2020	17:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111540

WorkList Name : fp101520 WorkList ID : 143332 Department : Wet-Chemistry Date : 10-15-2020 09:03:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/21/2020	Water	L4392-18	Flash Point	Cool 4 deg C	PSEG01	O12	HACK-MH	10/14/2020	1010A
10/16/2020	Water	L4403-01	Flash Point	Cool 4 deg C	TULL02	P11	SHELL-L15-L14-NORTH	10/14/2020	1010A
10/16/2020	Water	L4403-05	Flash Point	Cool 4 deg C	TULL02	P11	MH-N	10/14/2020	1010A
10/22/2020	Water	L4411-04	Flash Point	Cool 4 deg C	PSEG01	O12	32697	10/15/2020	1010A
10/22/2020	Water	L4418-02	Flash Point	Cool 4 deg C	ARCA05	P12	PB-2020-WW-000-02	10/14/2020	1010A

Date/Time 10/15/2020 14:45
 Received by: KS (Wetchem)
 Relinquished by: RECCGxt (ab)

Date/Time 10/15/2020 17:50
 Received by: RECCGxt (ab)
 Relinquished by: KS (Wetchem)