


LB111545

7470A

INSTRUMENT ID: CV4

Linear



A= 0.0000e+000

B= 2.6512e-004

C= 2.6353e-002

Rho= 0.9996439

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	25	0.000	25				
0.2	0.200	0.228	0.028	759	0.0 %	759				
2.5	2.500	2.551	0.051	9524	0.0 %	9524				
5.0	5.000	4.782	-0.218	17936	0.0 %	17936				
7.5	7.500	7.566	0.066	28438	0.0 %	28438				
10.0	10.000	10.041	0.041	37772	0.0 %	37772				

LB111545 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Oct 2020 09:03:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	7470A	PPB	16 Oct 2020 09:03:54	S	Std
0.2 - 1	50.2	759	-	0.2000	7470A	PPB	16 Oct 2020 09:06:12	S	Std
2.5 - 1	52.5	9524	-	2.5000	7470A	PPB	16 Oct 2020 09:08:29	S	Std
5.0 - 1	55	17936	-	5.0000	7470A	PPB	16 Oct 2020 09:10:46	S	Std
7.5 - 1	57.5	28438	-	7.5000	7470A	PPB	16 Oct 2020 09:13:02	S	Std
10.0 - 1	510	37772	-	10.0000	7470A	PPB	16 Oct 2020 09:15:18	S	Std
ICV59 - 1	ICV59	14084	3.7603	-	7470A	PPB	16 Oct 2020 09:20:35	U	SMPL
ICB59 - 1	ICB59	-57	0.0112	-	7470A	PPB	16 Oct 2020 09:22:50	U	SMPL
CCV38 - 1	CCV38	17798	4.7450	-	7470A	PPB	16 Oct 2020 09:25:05	U	SMPL
CCB38 - 1	CCB38	-79	0.0054	-	7470A	PPB	16 Oct 2020 09:27:20	U	SMPL
CRA - 1	CRA	709	0.2143	-	7470A	PPB	16 Oct 2020 09:29:35	U	SMPL
HighStd - 1	HighStd	37170	9.8810	-	7470A	PPB	16 Oct 2020 09:31:50	U	SMPL
ChkStd - 1	ChkStd	25054	6.6687	-	7470A	PPB	16 Oct 2020 09:34:05	U	SMPL
PB132408BL - 1	PBW	-178	-0.0208	-	7470A	PPB	16 Oct 2020 09:36:22	U	SMPL
PB132408BS - 1	LCSW	13344	3.5641	-	7470A	PPB	16 Oct 2020 09:38:38	U	SMPL
L4214-09 - 1	MDL-WATER-03-QT4-2020	96	0.0518	-	7470A	PPB	16 Oct 2020 09:46:16	U	SMPL
L4415-01 - 1	CALTANK1-20201014	29	0.0340	-	7470A	PPB	16 Oct 2020 09:48:32	U	SMPL
L4418-02 - 1	PB-2020-WW-000-02	457	0.1475	-	7470A	PPB	16 Oct 2020 09:50:48	U	SMPL
L4418-02DUP - 1	PB-2020-WW-000-02DUP	476	0.1526	-	7470A	PPB	16 Oct 2020 09:56:09	U	SMPL
L4418-02MS - 1	PB-2020-WW-000-02MS	14931	3.9849	-	7470A	PPB	16 Oct 2020 09:58:24	U	SMPL
CCV39 - 1	CCV39	17143	4.5713	-	7470A	PPB	16 Oct 2020 10:00:39	U	SMPL
CCB39 - 1	CCB39	-51	0.0128	-	7470A	PPB	16 Oct 2020 10:02:55	U	SMPL
L4418-02MSD - 1	PB-2020-WW-000-02MSD	15198	4.0557	-	7470A	PPB	16 Oct 2020 10:05:11	U	SMPL
L4392-17 - 1	HACK-MH	-55	0.0118	-	7470A	PPB	16 Oct 2020 10:07:27	U	SMPL
L4418-02LX5 - 1		108	0.0550	-	7470A	PPB	16 Oct 2020 10:09:42	U	SMPL
L4418-02A - 1		14708	3.9258	-	7470A	PPB	16 Oct 2020 10:11:58	U	SMPL
CCV40 - 1	CCV40	18728	4.9916	-	7470A	PPB	16 Oct 2020 10:14:14	U	SMPL
CCB40 - 1	CCB40	-56	0.0115	-	7470A	PPB	16 Oct 2020 10:16:30	U	SMPL