



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111547

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/15/2020	17:00
2	CAL2	1	Water	NA	NA	20.2	6.99	10/15/2020	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	10/15/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	10/15/2020	17:05
5	CCV1	1	Water	NA	NA	20.2	2.02	10/15/2020	17:06
6	L4412-07	1	Water	NA	NA	20.2	6.82	10/15/2020	17:09
7	L4412-07DUP	1	Water	NA	NA	20.3	6.83	10/15/2020	17:10
8	CCV2	1	Water	NA	NA	20.1	12.02	10/15/2020	17:11

WB 111547

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COROSIVITY L4412 WorkList ID : 143388 Department : Wet-Chemistry Date : 10-15-2020 16:42:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Water	L4412-07	Corrosivity	Cool 4 deg C	PSEG01	O12	98-02	10/15/2020	9040C

Date/Time 10-15-2020 11:50
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 10-15-2020 17:15
Received by: [Signature]
Relinquished by: [Signature]